



Illumination profile characterization of a light device for the dosimetry of intra-pleural photodynamic therapy for mesothelioma



Camille Munck^{a,b}, Serge Mordon^a, Nacim Betrouni^{a,*}

^a Univ. Lille, Inserm, CHU Lille, U1189 – ONCO-THAI – Image Assisted Laser Therapy for Oncology, F-59000 Lille, France

^b Department of Pulmonary and Thoracic Oncology, Lille University Hospital, Lille, France

ARTICLE INFO

Article history:

Received 25 May 2016

Received in revised form 27 July 2016

Accepted 31 August 2016

Available online 1 September 2016

Keywords:

Intrapleural photodynamic therapy

Dosimetry

Illumination profile

Light device

ABSTRACT

Background: Complete and homogeneous illumination of the pleural cavity is essential to the success of photodynamic therapy (PDT) for mesothelioma, but remains a challenge. Knowing the repartition and propagation of light around the light applicator could be the first step towards optimizing dosimetry. Here we propose a characterization method of the illumination profile of a specific light device.

Methods: The light wand, made of a cylindrical diffuser located inside an endotracheal tube, was fixed in a tank filled with dilute 0.01% intralipid. Light dosimetry was performed around the tip of the wand using two complementary methods: direct measurements of light power with an isotropic probe and measurements of light distribution characterization.

Results: Dosimetry with the isotropic probe showed an ellipse-shaped illumination. An optimized effective attenuation coefficient was deduced. Combined with the spatial representation, a theoretical illumination profile was established with iso-surfaces of fluence rate, defining a gradient light dose according to the distance from the diffuser.

Conclusion: A theoretical illumination profile of a light device was established and could be part of an intra-operative dosimetry system to improve light delivery during intrapleural PDT.

© 2016 Elsevier B.V. All rights reserved.

1. Introduction

Malignant pleural mesothelioma (MPM) is the primitive cancer of the pleura, occurring after exposure to asbestos fibers. It is a rare cancer, but its incidence is destined to rise throughout the world in the following decades, because of the increasing use of asbestos until the 1970's and since it is still used today in emerging countries. Its prognosis is poor with a median survival of less than a year, mainly due to complex treatments with disappointing outcomes. As part of a multimodal treatment, surgery offers the best chance at survival. However, even after the most complete tumor resection, microscopic tumor cells remain and it is essential to combine surgery with an efficient local adjuvant treatment. There are two main types of surgery: pleurectomy/decortication, where only the pleura is removed, and extra-pleural pneumonectomy, where the whole lung and the pleura are taken out. Whereas adjuvant hemithoracic radiotherapy or per operative intra-pleural therapies (chemotherapy, immunotherapy or gene therapy) have not proved

to be satisfactory in terms of survival, intra-pleural photodynamic therapy (PDT) has recently become a valid candidate.

Friedberg et al. (Philadelphia, USA) have evaluated the impact of intra-pleural PDT after surgical resection of MPM [1,2] and have demonstrated that adjuvant PDT improved the median survival significantly when associated with pleurectomy/decortication with a median survival was 31.7 months (95% [CI] 9–54.3 months) [1]. Their PDT protocol was the following [3]: patients were administered a photosensitizer, Photofrin® (porfimer sodium), 48 h before pleurectomy/decortications (P/D). Illumination of the pleural cavity was performed after maximal tumor resection, with a cylindrical light diffuser fixed inside an endotracheal tube and connected to a 635 nm laser source. To increase light diffusion, the pleural cavity and the cuff of the endotracheal tube were filled with a dilute intralipid solution of 0.01%. The light dose to deliver was 60 J/cm².

Illumination and dosimetry are key factors for a successful intra-pleural PDT, especially since the pleural cavity has a complex geometry. Knowing that, it seemed important to go back to the fundamentals of that light delivery and evaluate the light deployed by the device used for this procedure. The objective of this study was to produce an optimized light wand and propose a method to characterize the illumination profile of this light wand, in the clin-

* Corresponding author at: INSERM U1189 Onco-Thai, 1, Avenue Oscar Lambret, 59000, Lille, France.

E-mail address: nacim.betrouni@inserm.fr (N. Betrouni).



Fig. 1. The optimized light wand designed within laboratory. The cuff is filled with dilute intralipid liquid.

ical conditions of PDT for MPM. Characterization is done in term of fluence rate (W/cm^2) measurements.

2. Material and methods

2.1. Light wand construction and setting

Based on the device built by Friedberg et al. [3] a standardized light wand was designed using an endotracheal tube for structure. In addition, to ensure reproducibility for this study, the goals were to straighten the endotracheal tube for a better handling, and to fix the diffusing part of the optical fiber straight within the endotracheal tube and centered in the cuff. This light wand was composed of an endotracheal tube with an interior diameter of 8.5 mm (Super Safetyclear, Ruschelit[®], Teleflex Inc.) in which a carbon tube (length: 280 mm, exterior diameter: 8 mm, interior diameter: 7 mm) was inserted, to harden the endotracheal tube and correct its curvature. A piece of plexiglass tubing was shaped to fix and hold straight the diffusing part of the optical fiber, inside the inflated cuff, to guarantee homogenous illumination. Light was delivered using a 40 mm-tip-long cylindrical light diffuser (Model RD, Medlight[®] S.A., Switzerland). The cuff of the endotracheal was filled with 21 ml of intralipid solution at the concentration of 0.01% [3]. Fig. 1 represents the light wand assembled.

2.2. Illumination profile characterization

In order to characterize the illumination profile of the wand, it was set up in a specific container. A plastic tank of black color limiting light reflection was filled with approximately 2300 ml of dilute intralipid solution 0.01%, to reproduce the illumination of the pleural cavity. The tip of the light wand was fixed horizontally in the tank through a cable gland and the optical fiber was connected to a medical diode laser $635 \pm 3 \text{ nm}$ (Ceralas[®], Biolitec).

Light distribution around the diffusing tip of the light wand was then characterized by two complementary methods: direct measurements of the applied light power using sensors and estimation of the light gradient dose using a digital photography. The first allows a precise but partial sampling of the measurements while the second permits spatially characterization of light distribution.

2.2.1. Fluence rate measurements

An isotropic probe (Model IP 85, Medlight[®] S.A., Switzerland), collecting light in a large solid angle, was placed in a plastic tube and connected to a hand-held optical power meter (Model 841-PE, Newport[®] Corporation, Irvine, CA, USA). The tube is used as a support to prevent any movement from the probe, and so for a better accuracy and reproducibility of the measurements. The sensor tip is located outside the tube and in direct contact with the liquid. The isotropic probe was fixed vertically and moved above the diffusing tip of the light wand, using a benchmark of millimetric precision with a vertical and horizontal degree of freedom (Fig. 2(a)). Two-dimensional measurements (nW) of light power were taken every 5 mm around the cuff of the light wand, as represented on Fig. 2(b) and (c), for a total of 124 measurement points. This dosimetry

method had the advantage of giving precise measurements of the light emitted. The measurements were done with a laser power output of 1W and an angle of 0° along the endotracheal tube axis.

To test the reproducibility of this dosimetry method, the measurements were repeated three times, under the same experimental condition as described above, named “standard condition”. Experiments were also performed rotating the endotracheal tube at an angle of 0° , 90° , 180° and 270° along the endotracheal tube axis.

Last experiments were done by changing the laser power. In addition to 1W, the following powers were considered: 0.5W, 1.5 W, 2 W, 2.5 W and 3W.

As the aim is to obtain light fluence rate measurements, the power values had to be converted into fluence rate values. A calibration factor was estimated using the following protocol: a flat sensor with 1 cm^2 detection area (Model PD300, OPHIR, Israel) was used. Values recorded with this device were assumed to be the closest value for the local fluence rate (W/cm^2).

For the same position and using a defined power value output from the laser, a measure was acquired using the flat sensor and another using the isotropic probe.

Different measurements were done by changing, either the positions inside the tank or the power applied by the laser.

The complete pairwise values were used in a correlation analysis to estimate the calibration factor (CalibFactor (cm^{-2})).

2.2.2. Fluence rates distribution characterization

The purpose is to obtain a complete description of the light distribution around the applicator tip. A digital camera (SPIRICON, USA) with a sensor appropriate for the laser wavelength. Installed above the tank containing the applicator. The photography was taken with a camera position pertaining to have the same field of view as on Fig. 2b. This field of view corresponds to a distance of 19 cm above the intralipid solution surface.

A Photography was acquired for each power value (0.5 W, 1.0W, 1.5 W, 2 W, 2.5 W and 3 W).

The image as it is, is just a relative representation of the light attenuation around the applicator tip, calibrations are needed to make the image meaningful. Thus, it was processed by a home-made software (Optical Profiler) in the following way. First, using a ruler placed in the picture, a spatial calibration was done to estimate the pixel size, then the color map image was converted into a gray scale image and a normalization was done by dividing all the intensity values of the image by the maximum value. This value corresponds to the fiber trace represented by the brightest straight line on the image.

The second calibration is a pixel values calibration. It is based on the spatial matching of the punctual measurements done using the isoprobe with the picture. To this end, for each power values used as output from the laser, the points measurement array of Fig. 2b is positioned on the corresponding picture to define values iso-surfaces.

This matching allowed defining for each gray level a corresponding power value (i.e. a fluence rate value after conversion using the calibration factor).

3. Results

3.1. Fluence rate measurement

The measurements of the light power around the applicator, made with the isotropic probes, showed ellipsoidal surfaces as on Fig. 3. The light power decreased exponentially in a radial direction from the center of the diffuser.

Download English Version:

<https://daneshyari.com/en/article/3820315>

Download Persian Version:

<https://daneshyari.com/article/3820315>

[Daneshyari.com](https://daneshyari.com)