

# Accepted Manuscript

Evaluation of commercial kits for purification of circulating free DNA

Russell J. Diefenbach , Jenny H. Lee , Richard F. Kefford ,  
Helen Rizos

PII: S2210-7762(18)30250-3  
DOI: <https://doi.org/10.1016/j.cancergen.2018.08.005>  
Reference: CGEN 8478

To appear in: *Cancer Genetics*

Received date: 3 May 2018  
Revised date: 27 July 2018  
Accepted date: 19 August 2018

Please cite this article as: Russell J. Diefenbach , Jenny H. Lee , Richard F. Kefford , Helen Rizos , Evaluation of commercial kits for purification of circulating free DNA, *Cancer Genetics* (2018), doi: <https://doi.org/10.1016/j.cancergen.2018.08.005>



This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

**Highlights**

- Spin column-based kits outperform magnetic bead-based commercial cfDNA kits
- The Qiagen spin column-based cfDNA kit remains the gold standard
- The Qiagen magnetic bead-based cfDNA kit is an alternative when factoring in price

ACCEPTED MANUSCRIPT

Download English Version:

<https://daneshyari.com/en/article/11007681>

Download Persian Version:

<https://daneshyari.com/article/11007681>

[Daneshyari.com](https://daneshyari.com)