



## Quantitative determination of cyclic polylactic acid oligomers in serum by direct injection liquid chromatography tandem mass spectrometry

Issey Osaka<sup>a</sup>, Arihumi Yoshimoto<sup>a</sup>, Mikio Watanabe<sup>b</sup>, Masashi Takama<sup>c</sup>, Masahiro Murakami<sup>c</sup>, Hideya Kawasaki<sup>a</sup>, Ryuichi Arakawa<sup>a,\*</sup>

<sup>a</sup> Department of Applied Chemistry, Kansai University, 3-3-35 Yamatecho, Suita, Osaka 564-8680, Japan

<sup>b</sup> Department of Chemistry, School of Science, Tokai University, 1117 Kitakaname Hiratuka-shi, Kanagawa 259-1292, Japan

<sup>c</sup> Laboratory of Pharmaceutics, Faculty of Pharmacy, Osaka Ohtani University, 3-11-1 Nisikiori-Kita, Tondabayasi, Osaka 584-8540, Japan

### ARTICLE INFO

#### Article history:

Received 29 February 2008

Accepted 22 June 2008

Available online 27 June 2008

#### Keywords:

Cyclic polylactic acid oligomer

LC–MS/MS

Deproteination

Direct injection of serum sample

BSA-coated column

### ABSTRACT

Poly(lactic acid) (PLA) is a biodegradable polymer, currently used in pharmaceutical and surgical devices. There is a concern that cyclic polylactic acid (CPLA), which is a by-product of PLA synthesis, may be introduced into the human body as an undesirable contaminant. We carried out a quantitation investigation of the CPLA heptamer (CPLA-7) by liquid chromatography mass spectrometry (LC–MS). We found that CPLA-7 binds strongly with serum proteins and that only 62% of CPLA-7 was recovered after routine deproteination; therefore, we directly injected serum into the LC–MS/MS system after passage through a bovine serum albumin (BSA)-coated chromatographic column and found the recovery of CPLA-7 was improved to 84%, and that the detection ( $S/N = 3$ ) and quantitation limit ( $S/N = 10$  and below 15% relative standard deviation) were 1.5 and 2.5 ng/mL, respectively. We conclude that direct injection LC–MS/MS, using a BSA column, is a simple and effective quantitative analysis method for CPLA in serum.

© 2008 Elsevier B.V. All rights reserved.

### 1. Introduction

Biodegradable polymers have been widely used as industrial and medical materials, and their degradation products are present throughout the environment [1]. It is possible for biodegradable polymers or their products to enter the human body not only via the environment, but also as medical materials in drug delivery and surgical devices, and even as drugs themselves. It is thus important to develop an effective quantification method for biodegradable polymers or oligomers in a biological matrix rather than in environmental samples.


Numerous studies have examined the degradation of biodegradable polymers [2–4], among which, size exclusion chromatography (SEC) [5–7], NMR [8], X-ray diffraction [8], and laser diffractometry [9] have been used for the identification and quantitation of these polymers. Recently, mass spectrometry and X-ray photoelectron spectrometry (XPS) [10–13] have also been used for structural analysis of biodegradable polymers. Studies on decomposition products at the surface of poly( $\beta$ -maleic acid) and polylactic acid (PLA) by secondary ion mass spectrometry (SIMS) and XPS have been reported [10–12]. Using TOF–SIMS, *in vitro* hydrolytic reactions of

PLA and lactic-co-glycolic acid were studied [14]. Matrix-assisted laser desorption ionization mass spectrometry (MALDI–MS) is a soft ionization technique suitable for the analysis of thermally-labile or high molecular weight molecules and is particularly effective for structural analysis of synthetic polymers; several analyses of polystyrene, polyesters, or their copolymers have been reported [15–18]. Using the melt polycondensation method, Kéki et al. investigated the temperature dependence on by-product formation of PLA synthesis from D,L-lactic acid [19]. Tandem mass spectrometry is an effective technique for detailed structural analysis; for example, cyclic polylactic acid (CPLA) and linear polylactic acid (LPLA) can be differentiated by their different fragmentations [20].

PLA is one of the biodegradable polymers used in drug delivery and surgical devices, or as a material for postoperative adhesion prevention. PLA decomposes to lactic acid which is bio-resorbable and is harmlessly absorbed by human body. MALDI–MS analysis has revealed that chemical synthesis of PLA results in the formation of LPLA as the principal product and CPLA as a by-product [19]. In addition, it was recently reported that CPLA can lower the activities of pyruvate kinase and lactic dehydrogenase, which may lead to the development of drugs suppressing FM3A ascites tumor cells [21]; therefore, it is important to develop an effective method of quantitatively analyzing CPLA in plasma or serum and to study its metabolism.

\* Corresponding author. Tel.: +81 6 6368 0781; fax: +81 6 6339 4026.

E-mail address: [arak@ipcku.kansai-u.ac.jp](mailto:arak@ipcku.kansai-u.ac.jp) (R. Arakawa).



**Fig. 1.** Structures of (a) cyclic poly(lactic acid) heptamer (CPLA-7) and (b) its deuterated isotope (CPLA-7-d<sub>3</sub>).

Gas chromatography mass spectrometry (GC–MS) and liquid chromatography mass spectrometry (LC–MS) are commonly used for highly sensitive quantitation of organic compounds in plasma or serum. Quantitative analysis of plasma lactic acid using GC–MS [22] and ion-exclusion chromatography [23] has been reported, but our literature search yielded no reports on quantitative analysis of LPLA or CPLA oligomers.

In the present study, we used lactic acid, lactoyllactic acid, and L-lactate-3,3,3-d<sub>3</sub> to synthesize a heptamer of CPLA (CPLA-7) and its isotope CPLA-7-d<sub>3</sub>. Due to strong binding between CPLA-7 and serum proteins, the LC–MS/MS analysis followed by regular deproteination resulted in low recovery of CPLA-7; therefore, we developed a method of directly injecting CPLA-7 contained in serum into LC–MS/MS, employing a bovine serum albumin (BSA)-coated column, and investigated the optimal conditions for this quantitation.

## 2. Experimental procedures

### 2.1. Synthesis of CPLA-7 and CPLA-7-d<sub>3</sub>

The O-silylated trimers of lactic acid benzylester were obtained when O-*tert*-butyldimethylsilylated lactoyllactic acid was allowed to react with benzylated lactoyllactic acid or lactic acid via the Keck protocol. After selective deprotection of hydroxyl or carboxyl groups of the products obtained, the condensation reactions of these products using dicyclohexylcarbodiimide (DCC) catalyzed 4-(dimethylamino)pyridine (DMAP) were carried out to yield O-silylated oligo lactic acid benzylester. The protective groups were removed to produce a free heptamer of lactic acid. Finally, this heptamer was allowed to react with 2,4,6-trichlorobenzoyl chloride in the presence of amine, followed by DMAP-catalyzed cyclization under a dilute condition (Yamaguchi method). The mixture was purified by silica gel column chromatography to obtain monomeric CPLA-7 (Fig. 1a).

Benzyl L-lactate-3,3,3-d<sub>3</sub> was synthesized by reacting the sodium salt of L-lactic acid-d<sub>3</sub> (>98 atom%, Isotec Inc.) with benzylbromide. The siloxylactate hexamer was synthesized by

condensation of linear lactate trimer benzylester and siloxylactate trimer benzylester. The siloxylactate hexamer was then condensed with benzyl L-lactate-3,3,3-d<sub>3</sub> to form linear lactate heptamer-d<sub>3</sub>, which was cyclized and purified as described in CPLA-7 synthesis to obtain monomeric CPLA-7-d<sub>3</sub> (Fig. 1b). The total yield of synthesis for CPLA-7-d<sub>3</sub> was ca. 50% just as for CPLA-7.

### 2.2. Reagents and standards

CPLA-7 and CPLA-7-d<sub>3</sub> were dissolved in acetonitrile/water (1:1, v/v) and diluted with water. A 100-ng/mL solution of CPLA-7-d<sub>3</sub> was used as the internal standard. Acetonitrile (LC–MS grade), phosphoric acid (HPLC grade) and ammonium acetate were purchased from WAKO Pure Chemicals (Osaka, Japan). Ultrapure water was obtained using a water distillation apparatus (RFD250NB, Advantec). Human serum from normal subjects was obtained from CHEMICON Int. Inc. (Temecula, CA, USA).

### 2.3. Sample pretreatment

For direct serum sample injection, sera containing 1% phosphoric acid, CPLA-7 solution and internal standard solution were mixed at a ratio of 2/1/1 (v/v/v). The solution was centrifuged (Centrifugeb 5417R, Eppendorf) for 3 min at 14,000 rpm, and the supernatant was injected directly into LC–MS/MS.

### 2.4. Apparatus

#### 2.4.1. Liquid chromatography

HPLC was performed using an Agilent HP1100. A BSA-coated chromatographic column, the BSA-ODS-100V (2.0 mm × 50 mm, 5 μm particle size), supplied by Tosoh Co., Japan was used and was found to effectively remove proteins in the sample. The principle of separation for this column is as follows. Target compounds penetrate silica pores of column particles and are retained by the ODS interaction in these pores. At the same time, high molecular weight proteins do not penetrate the pores and are not retained. Furthermore, the surfaces of the particles are coated with polar BSA so that serum protein is readily eluted from the column, without retention or denaturalization. Binary mobile phases A (H<sub>2</sub>O) and B (95% acetonitrile) were used with a flow rate of 0.3 mL/min and a gradient elution: programmed from 0% B (0–3.5 min) to 80% B (3.5–6.5 min), held at 80% B (6.5–8.5 min), returned to 0% B (8.5–9.0 min), and washed with 0% B (9–12 min). The injection volume was 5 μL. A cationizing solution (50 mM ammonium carbonate) with a flow rate of 0.06 mL/min was added at the post-column stage. Using a 6-way switch valve, flow between 7.5 and 9.5 min was introduced into the mass spectrometer and the flows before and after this time period were discarded as waste.

#### 2.4.2. ESI–MS/MS

ESI mass spectra were obtained by a triple quadrupole mass spectrometer TSQ (ThermoQuest) in positive ion mode. The capillary temperature was 200 °C, spray voltage 4.5 kV, auxiliary gas 20 psi, and sheath gas 60 psi. The collision energy of the parent ion was set at 30 eV.

## 3. Results and discussion

### 3.1. Optimization of ESI–MS/MS

The ESI spectrum of CPLA-7 in Fig. 2 exhibits three molecular adduct ions at *m/z* 522, 527 and 543, corresponding to [M+NH<sub>4</sub>]<sup>+</sup>,

Download English Version:

<https://daneshyari.com/en/article/1216776>

Download Persian Version:

<https://daneshyari.com/article/1216776>

[Daneshyari.com](https://daneshyari.com)