

Short communication

Modified high-performance liquid chromatography with electrochemical detection method for plasma measurement of levodopa, 3-*O*-methyldopa, dopamine, carbidopa and 3,4-dihydroxyphenyl acetic acidMorvarid Karimi^a, Juanita L. Carl^{a,b}, Susan Loftin^{a,b}, Joel S. Perlmutter^{a,b,c,d,*}^a Department of Neurology, Washington University in St. Louis, St. Louis, MO, USA^b Mallinckrodt Institute of Radiology, Washington University in St. Louis, St. Louis, MO, USA^c Program in Physical Therapy, Washington University in St. Louis, St. Louis, MO, USA^d Department of Anatomy & Neurobiology, Washington University in St. Louis, St. Louis, MO, USA

Received 23 December 2005; accepted 10 March 2006

Available online 11 April 2006

Abstract

Plasma measurements of levodopa and its major metabolites including dopamine and 3-*O*-methyldopa have been limited by cumbersome methods and poor sensitivity within relatively narrow ranges of plasma levels. We now report a modification of an HPLC method that permits concomitant measurements of a wide range of concentrations of levodopa, dopamine (DA), carbidopa, 3-*O*-methyldopa (3-OMD) and 3,4-dihydroxyphenyl acetic acid (DOPAC) from one HPLC injection. The recoveries ranged from 77 to 107% with an intra-day precision around 5% (CV) and inter-day CV's about 10–20%. This validated method will simplify pharmacokinetic studies of levodopa and its metabolites for mechanistic studies or therapeutic clinical monitoring which play a crucial role in development of strategies to prolong motor benefits from individual doses and reduce involuntary movements called dyskinesias.

© 2006 Elsevier B.V. All rights reserved.

Keywords: Levodopa; Plasma measurements; HPLC; Carbidopa; 3-*O*-Methyldopa; Dopamine; DOPAC (3,4-dihydroxyphenyl acetic acid)

1. Introduction

Abnormalities of cerebral dopaminergic pathways play an important role in the pathophysiology or treatment of several neuropsychiatric diseases including Parkinson disease (PD). Degeneration of nigrostriatal neurons with subsequent loss of striatal dopamine produces many of the clinical manifestations of PD [1] and pharmacotherapy with levodopa, the immediate precursor of dopamine, provides substantial clinical benefit [2]. Concomitant administration of a peripheral decarboxylase inhibitor like carbidopa that does not cross the blood brain barrier reduces systemic production of dopamine and reduces untoward effects such as nausea [3]. However, as the disease progresses and treatment continues many patients with PD develop other drug related side effects such as short duration of response to

individual doses and involuntary movements called dyskinesias [4]. Inhibitors of catechol-*O*-methyl transferase (COMT) reduce *O*-methylation of levodopa and dopamine thereby potentially prolong motor benefit from individual doses of levodopa [5]. Measurements of levodopa and its metabolites in blood have been crucial for development of these strategies by clarifying the role of systemic pharmacokinetics in the clinical responses [6].

Plasma measurements of levodopa, 3-*O*-methyldopa (3-OMD), dopamine (DA) and carbidopa, however, have been technically difficult in people with PD treated with pharmacologic doses of levodopa in the presence of a decarboxylase inhibitor since levodopa and 3-*O*-methyldopa levels tend to be high, at least 1000-fold higher than plasma dopamine [7]. Measurements of one of the metabolites of dopamine 3,4-dihydroxyphenyl acetic acid (DOPAC) also is useful since DOPAC may be involved in the pathogenesis of PD. Recent animal studies found that DOPAC reacts with nitric oxide in mitochondria to produce compounds that inhibit O₂ uptake that could lead to cell death [8]. Rondelli et al. [9] developed a high performance

* Correspondence to: 4525 Scott Avenue, Campus Box 8225, St. Louis, MO 63110, USA. Tel.: +1 314 362 6908; fax: +1 314 362 0168.

E-mail address: joel@npg.wustl.edu (J.S. Perlmutter).

liquid chromatography (HPLC)-based method for levodopa, 3-*O*-methyldopa dopamine and levodopa methyl ester. We now report a modification of this method that permits simultaneous measurements of carbidopa and DOPAC in addition to other analytes.

2. Methods and materials

2.1. Subjects

We took blood samples from eight patients (seven males) with clinical diagnoses of PD. Their age ranged from 55 to 83 years with an average of 70. They had a weight range of 45–93 kg with an average of 76.5 kg. Four of them had been treated chronically with levodopa but refrained from taking medication overnight prior to these studies. Each subject was given carbidopa 200 mg orally at least 1 h prior to oral intake of levodopa/carbidopa 150 mg/37.5 mg. Three milliliter samples were drawn from the intravenous line for 3 h after levodopa. These studies were approved by the Human Studies Committee of Washington University in St. Louis, and all subjects gave written, informed consent.

2.2. Materials and reagents

L-DOPA, DOPAC, dopamine, and 4-dihydroxybenzylamine (DHBA internal standard) were purchased from RBI-Sigma, USP grade. Carbidopa (purity >98%), 3-OMD (purity >99%), heparin, EGTA, glutathione, sodium metabisulfite and perchloric acid were obtained from Sigma–Aldrich, St. Louis, MO. Acetonitrile and methanol were of HPLC grade and obtained from Fisher Scientific. CAT-A-PHASE HPLC buffer, designed for analysis of catecholamines in plasma, was from ESA (catalog #45-0180, Chelmsford, MA). The buffer contains methanol, phosphate buffer and a patented ion pairing agent. We modified the CAT-A-PHASE HPLC buffer by addition of 0.3% acetonitrile and adjusted pH to 3.20 ± 0.03 with 2 N sodium hydroxide to facilitate optimal separation of all compounds. The pH adjustment was checked before every use, as small pH changes produce large shifts in retention time for levodopa, 3-*O*-methyldopa and carbidopa.

2.3. Standard solutions

Standards were prepared freshly for each sample run using 1 mg/ml stock solutions stored frozen at –80 °C (0.1 M perchloric acid (PCA+) containing 0.1% sodium metabisulfite). A standard mixture of 250 µg levodopa, 500 µg 3-*O*-methyldopa, 200 µg carbidopa, 25 µg dopamine, and 25 µg DOPAC in 1 ml was prepared using aliquots of the stock standards and serially diluted as above to give concentrations that ranged across two orders of magnitude in a total of six dilutions. Each final standard sample contained 0.5 ml aliquot of plasma from a normal not treated with levodopa, 10 µl (0.5 µg) internal standard (1 mg/ml DHBA diluted to 50 µg/ml in PCA+) and 10 µl aliquot of one of the dilutions to yield six standard samples. The standard curve was obtained by calculating the peak–height ratios

of each compound to the internal standard plotted against the known concentration of each substance.

2.4. HPLC

We used a Coulochem II HPLC system (ESA, Inc.) with an ESA Model 580 Solvent Delivery Module, a Rheodyne 9125 Sample Injector, and an ESA catecholamine HR-80 Column (8 cm × 4.6 mm I.D. packed with 3-µm C₁₈ stationary phase) protected by a Newguard Cartridge (15 mm × 3.2 mm I.D. packed with 7-µm spherical RP-18 Aquapore, a Brownlee product). The conditioning cell was ESA Model 5021 and the analytical cell was ESA model 5011. The system operated in redox mode. The optimal selection of the applied electrode potentials were determined by generating current–voltage curves for each compound. The conditioning cell was set at 450 mV. The first electrode in the analytical cell was set at 50 mV, and the second one at –400 mV. Plasma levels of levodopa and frequently 3-OMD (especially in chronically treated patients) are 100–1000 higher than DOPAC, dopamine and carbidopa. We optimized detector settings depending on the compound with a full-scale gain varying from 100 nA to 2 mA during the run. The change to the more sensitive gain was done before or after the internal standard peak. Output from the second analytical cell was recorded on a ChromJet integrator, Model SP4400 (Thermoseparation Products, Fremont CA, USA). A flow rate of 0.7 ml/min was used, and separations were run at room temperature (about 21 °C). The mobile phase was not recycled during the day of analysis. The length of analysis was 20 min. Up to 20 runs including standards and samples were performed over 8–10 h.

2.5. Sample preparation

Blood samples were collected in a syringe and immediately transferred to 6 ml polypropylene tubes containing 5.7 mg EGTA, 3.6 mg reduced glutathione, and 50 units heparin in 60 µl on ice to maintain stability of the plasma catecholamines. This solution was prepared fresh every 14 days with the pH adjusted to 6.5–7.0. The blood samples remained on ice less than 30 min until centrifuged. Whole blood was spun at 1250 × *g* for 30 min in a CRU-5000 centrifuge IEC at 4 °C. The plasma was transferred to 6 ml polypropylene tubes, snap frozen on dry ice, and stored in a –80 °C freezer until analyzed. Samples were thawed, extracted, and analyzed on the same day. Plasma (0.5 ml) obtained from a normal subject not exposed to levodopa used as a blank, and plasma samples from patients were spiked with 10 µl (0.5 µg) internal standard (1 mg/ml DHBA diluted to 50 µg/ml in PCA+). To precipitate proteins perchloric acid (1.2 M) 0.3 ml was added to the plasma. Tubes were mixed, stored on ice for 10 min and spun for 7 min at 1250 × *g* at 4 °C; 300 µl of the supernatant were added to 200 µl of 2 M potassium citrate buffer (pH 3.8) to precipitate the perchlorate. Each tube was vortexed for 1 min, left on ice for 10 min, then centrifuged as above. Three hundred microliters of the clear supernatant were filtered through 0.22 mm syringe filters and 20 µl of the supernatant were injected onto the HPLC for analysis. Recov-

Download English Version:

<https://daneshyari.com/en/article/1217374>

Download Persian Version:

<https://daneshyari.com/article/1217374>

[Daneshyari.com](https://daneshyari.com)