

Fluorescence probe enhanced spectrofluorimetric method for the determination of gatifloxacin in pharmaceutical formulations and biological fluids

Xiashi Zhu*, Aiqin Gong, Suhai Yu

College of Chemistry & Chemical Engineering, Yangzhou University, Yangzhou 225002, China

Received 16 July 2006; accepted 25 April 2007

Abstract

A spectrofluorimetry for the determination of gatifloxacin (GFLX) was developed based on the strong fluorescence of gatifloxacin after adding fluorescence probe yttrium in buffer solution (pH 7.0) and various factors of influencing fluorescence have been researched. Under the optimum conditions, the liner range was 4.00×10^{-8} to 1.00×10^{-6} g mL⁻¹ and the detection limit is 3.36×10^{-9} g mL⁻¹ (correlation coefficient $r = 0.9997$), respectively. The relative standard deviation was 1.1% for 11 measurements of 5.6×10^{-7} g mL⁻¹ gatifloxacin standard solution. The mechanism of sensitizing effect of probe was discussed. The proposed method has been successfully applied to determine real samples and the obtained results are in good agreement with the results of HPLC.

© 2007 Published by Elsevier B.V.

Keywords: Gatifloxacin; Spectrofluorimetry; Probe

1. Introduction

Gatifloxacin { \pm -1-cyclopropyl-6-fluoro-7-(3-methyl-1-piperaziny)-8-methoxy-1,4-dihydro-4-oxo-3-quinolinecarboxylic acid sesquihydrate} is a new broad-spectrum 8-methoxy fluoroquinolone antibacterial agent. Its molecular formula is C₁₉H₂₂FN₃O₄·1.5H₂O and its molecular weight is 402.23. The structure of GFLX is shown in Fig. 1.

The new fluoroquinolone, which has shown activity against both Gram (+), Gram (–) and anaerobic species, behaves excellent pharmacokinetic and pharmacodynamic characters. Its negative action is very small and even no photoactive reaction [1–3]. Thus, a rapid, low cost, and accessible procedure for the quantification of gatifloxacin in dosage forms is desirable. Various techniques have been utilized for the determination of the fluoroquinolone, among them HPLC is the most frequently used technique [4,5]. Other analytical methods are spectrophotometry (UV) [6], capillary electrophoresis [7] and fluorospectrophotometry [1,3,8], FIA-chemiluminescence [9],

etc. These methods were limited to widely use because of expensive instruments cost or low sensitivity.

Probes are some small organics, inorganic ions, metal complex ions and quantum points. The spectrophotometric or electrochemical characters of the analytes would change when they form extra-molecular complexes with probes, therefore it could provide concentrative or structural information of the analytes or enhance sensitivity of determination [10]. Nowadays there are many cases to use probes in assaying pharmaceuticals. Literature [11] studied the reaction of oxyfloxacin with DNA using Tb as probe. There was also a paper [12] that used methylene blue as probe to research the reaction of *p*-(*N,N*-dimethyl aminomethyl) calix [8] arene with DNA. Some papers have reported to determine gatifloxacin by spectrofluorimetry without probes [1,8], but the sensitivity was low. The sensitivity of assay gatifloxacin could be enhanced in the medium of SDS micelle solution [3]. Fluorescence probe enhanced spectrofluorimetric method for the determination of GFLX seems to be lacking. In this paper, Y³⁺ as probe, spectrofluorimetric method for the determination of GFLX is investigated. To the best of our knowledge, this is the first attempt to determine GFLX in pharmaceutical formulations by fluorescence probe enhanced spectrofluorimetry. The main advantages of the proposed

* Corresponding author. Tel.: +86 514 7975244; fax: +86 514 7975244.
E-mail addresses: xsazu@yzu.edu.cn, zhuxiashi@sina.com (X. Zhu).




Fig. 1. The structure of gatifloxacin.

methods compared with previously reported are their simplicity and lower cost. The detection limits are comparable or better than the detection limits of the other methods [1,3,8].

2. Experimental

2.1. Chemicals

Gatifloxacin was kindly provided by Jinghua institution. Gatifloxacin capsule and injection were from Biocause (Hunai, China) and Suzhong (Jiangsu, China) pharmaceutical industries, respectively. Acetic acid, ammonium acetate, hydrochloric acid and Y_2O_3 were purchased from reagent company of Guoyao (Shanghai, China). All chemicals were of analytical reagent grade. Water was distilled, deionised.

Gatifloxacin stock standard solution of $2.0 \times 10^{-4} \text{ g mL}^{-1}$ was prepared by dissolving 0.0100 g of gatifloxacin in 50 mL of 0.1 mol L^{-1} HCl and kept in the dark. Working standard of $2.0 \times 10^{-5} \text{ g mL}^{-1}$ was prepared daily by dilution of stock standard solution with high-purity water. Buffer solution (pH 7.0) was prepared by dissolving 77.0 g ammonium acetate in 500 mL high purity water and was adjusted by the addition of acetate acid. The solution of Y^{3+} was prepared by dissolving 0.1270 g Y_2O_3 in 1 mL of HCl and diluted to 100 mL with water.

2.2. Apparatus

An Hitachi F-4500 spectrofluorimeter (Japan) was used for all the measurement, with excitation and emission slits at 2.5 nm, $\lambda_{\text{exc}} = 292 \text{ nm}$ and 1 cm quartz cell. The pH was measured on a pH S-25 pH-meter (Shanghai).

2.3. Sample preparation

The solution of gatifloxacin capsule and gatifloxacin injection were prepared by dissolving suitable amount of the commercial samples in 0.1 mol L^{-1} HCl and diluted the resulting solution to adjust the concentration with water to that required by the experimental conditions adopted.

A 1.0 mL serum sample was deproteinized by adding 4.0 mL 10% trichloroacetic acid (CCl_3COOH) in a centrifuge tube, which was then centrifuged for 15 min at 4000 rpm. The supernatant was diluted with deionized water to suitable concentration. No further pre-treatment was required for urine samples. The

treated serum and urine samples were spiked with convenient amounts of GFLX stock solutions. Human serum and urine were kindly provided by healthy volunteers.

2.4. Determination of gatifloxacin

Aliquots of working solutions or sample solutions of gatifloxacin, 3.0 mL of acetic acid–ammonium acetate buffer (pH 7.0), 1.0 mL of Y_2O_3 were pipetted in to 25 mL calibrated flasks and diluted to the mark with water. The obtained solutions were thermostated at $20.0 \pm 0.5^\circ\text{C}$ and the fluorescence intensity was measured, using an excitation wavelength of 292 nm, against a blank solution.

2.5. Determination of relative fluorescence quantum yields

Fluorescence quantum yields of GFLX and $GFLX-Y^{3+}$ were measured using $1.0 \times 10^{-6} \text{ g mL}^{-1}$ quinine sulfate as reference material. Under the same apparatus conditions, according to the equation $\varphi_2 = (\varphi_1 A_1 F_2) / (F_1 A_2)$ [13], the quantum yields of the analyte was calculated. Briefly, φ_1 and φ_2 are corresponding the standard and unknown fluorescence quantum yield and F_1 and F_2 the integral areas of two calibration fluorescence emission curves, A_1 and A_2 the absorbance ($\lambda_{\text{absorbance}} = \lambda_{\text{emission}}$) of the standard and unknown and $\varphi_1 = 0.55$ (25°C) is known.

3. Results and discussion

3.1. The fluorescence spectra of gatifloxacin

The fluorescence spectra of gatifloxacin in pH 7.0 buffer solution are shown in Fig. 2. Comparing the curves, it was found the fluorescence intensity was enhanced about one time when using $4.0 \times 10^{-5} \text{ g mL}^{-1} Y^{3+}$ as probe.

3.2. The selection of probes

The selection of probes is very important in spectrofluorimetry. In this work, probes such as aluminum, terbium, europium,




Fig. 2. Fluorescence spectra of GFLX (GFLX: $4.0 \times 10^{-7} \text{ g mL}^{-1}$).

Download English Version:

<https://daneshyari.com/en/article/1235313>

Download Persian Version:

<https://daneshyari.com/article/1235313>

[Daneshyari.com](https://daneshyari.com)