



DIABETIC RETINOPATHY UPDATE

# Microperimetry in diabetic retinopathy

Edoardo Midena, MD, PhD <sup>a,b,\*</sup>, Stela Vujosevic, MD <sup>b</sup>

<sup>a</sup> Department of Ophthalmology, University of Padova, Padova, Italy

<sup>b</sup> Fondazione GB Bietti per l'Oftalmologia, IRCCS, Roma, Italy

Received 18 January 2011; accepted 21 January 2011

Available online 31 January 2011

**KEYWORDS**

Diabetic retinopathy;  
Visual function;  
Microperimetry

**Abstract** Diabetic retinopathy has an enormous impact on visual function, even before permanent visual acuity loss. Moreover, adequate functional tests are mandatory to diagnose and follow diabetic patients treated for diabetic macular edema (DME). More precisely, the visual function safety profile of any therapy for DME should be accurately investigated. Microperimetry offers the possibility to obtain an exact fundus-related quantification of retinal sensitivity, and it is changing the current approach to the functional investigation of diabetic retinopathy.

© 2011 King Saud University. Production and hosting by Elsevier B.V. All rights reserved.

**Contents**

1. Introduction . . . . .	131
2. Microperimetry (fundus-related perimetry). . . . .	132
3. Conclusion . . . . .	134
References . . . . .	134


\* Corresponding author at: Department of Ophthalmology, Via Giustiniani 2, 35128 Padova, Italy.

E-mail address: edoardo.midena@unipd.it (E. Midena).

1319-4534 © 2011 King Saud University. Production and hosting by Elsevier B.V. All rights reserved.

Peer review under responsibility of King Saud University.

doi:10.1016/j.sjopt.2011.01.010



## 1. Introduction

Diabetic retinopathy is one of the major causes of permanent visual (acuity) loss in the working population. Moreover, the prevalence of diabetes mellitus is dramatically increasing worldwide. Visual loss is commonly quantified by a full contrast visual acuity test (by Snellen or ETDRS charts). Unfortunately, this full-contrast visual acuity test doesn't reflect the real visual functional abnormalities due to the retinal involvement secondary to diabetes mellitus. Moreover, subtle and precocious neurosensory visual abnormalities have been quantified in diabetic patients in order to detect early visual dysfunction, even before the onset of clinically detectable

retinopathy. The aim of these investigations is to try to identify among diabetic subjects a population at higher risk of developing vision threatening retinopathy (Bresnick, 1986; Midena et al., 1990). Psychophysical visual function testing may reflect the neural activity of the whole visual pathway, but it is known that psychophysical tests are valuable clinical indicators of retinal function derangements induced by the metabolic changes secondary to diabetes mellitus. In fact, in diabetic patients impaired vision in dim light and difficulties in recognizing the contour of objects in low contrast conditions are common complaints even with good visual acuity and full visual fields (Hyvärinen et al., 1983). Visual acuity is still considered the gold standard in clinical practice of vision testing, but it does not entirely reflect functional vision. Functional vision describes the impact of sight on the quality of life that represents *the patient's point of view* (Sharma et al., 2005; Owsley and Sloane, 1987; Midena, 2006). This approach is better quantified using all available psychophysical tests, mainly analyzing fundus-related retinal sensitivity threshold. This paper reviews the current application of fundus-related perimetry, better known as microperimetry, in the diagnosis and follow-up of diabetic retinopathy.

## 2. Microperimetry (fundus-related perimetry)

Perimetry encompasses the assessment of differential light threshold of retinal locations from the fovea to the pre-planned periphery. Static perimetry is particularly useful for detailed probing in carefully selected areas and represents the current cornerstone of visual field testing. Standard threshold static automated perimetry quantifies the differential light threshold required to detect a static white light stimulus in the visual field. Since standard threshold perimetry uses a static achromatic stimulus, it is thought to non-selectively evoke both major groups of retinal ganglion cells. Newer technologies are aimed at earlier detection of subtle deficits and enhancing diagnostic accuracy. In diabetic macular edema (DME), visual acuity loss is quite relevant and irreversible when long lasting edema involves the center of the macula; in these cases the outcome of laser treatment is poor. But, before the loss of visual acuity is reported by patients, they may suffer from other disturbances of visual function such as: waviness, blurring, relative scotoma and decrease of contrast sensitivity which are not assessed and quantified in routine examination. Therefore, a visual function test aimed at identifying vision threatening retinopathy before visual acuity is affected would be of great value. One possible approach may be to identify decreased sensitivity in central and paracentral areas using microperimetry (Midena, 2006). As elegantly stated by Sunnes et al., conventional visual field examination is inadequate for the accurate functional evaluation of macular diseases and detection of small scotoma, particularly when foveal function is compromised and the patient may have unstable and extrafoveal fixation (Sunnes et al., 1995). Accuracy of the conventional visual field rests on the assumption that fixation is foveal and stable. Moreover, the detection of the site and stability of retinal fixation (foveal or extrafoveal) and the quantification of retinal threshold over small and discrete retinal lesions are beyond the possibilities of conventional, automatic and non automatic perimetry (Midena and Radin, 2006). The integration of retinal details with function has been achieved

by fundus-related perimetry, more widely known as microperimetry. Microperimetry allows for the exact topographic correlation between fundus abnormalities and corresponding functional alterations by integration, with different methods, of differential light threshold (more commonly known as retinal sensitivity) and fundus imaging. It also allows to quantify fixation characteristics, by exactly defining location and stability of any foveal or extrafoveal (PRL: preferred retinal locus) fixation site, as well as determination of size, site and shape of scotoma. Moreover, the possibility of an automatic follow-up examination (using the microperimeter MP1, Nidek Co, Japan) which allows the evaluation of exactly the same retinal points tested at the baseline, regardless any change in fixation characteristics is a valuable tool of this technique, mainly in the evaluation of treatment outcome. Microperimetry offers several advantages versus standard perimetry in the quantification of macular sensitivity, such as: direct real time fundus control; direct correlation between sensitivity and fundus details; detection of central microscotomata; continuous monitoring of fixation. The original Scanning Laser Ophthalmoscope (SLO, Rodenstock, Germany) was the first instrument combining static perimetric testing and simultaneous observation of the fundus. SLO allowed a real-time examination by an infrared (IR) source of the retina and allowed the manual projection of visual stimuli of different shapes, sizes and intensities over selected retinal areas. The sensitivity map, obtained according to the stimulation pattern (in dB or pseudocolors), was available at the end of the examination. This map contained the fixation area, the fixation target, and the threshold data. This instrument is no more commercially available.

With the introduction of a new microperimeter, a liquid crystal display (LCD) microperimeter (MP1) with a coupled color fundus camera, visualization of color fundus details allows to directly report functional data onto clinical fundus image and automatic tests are also obtained. MP-1 microperimeter has both an infrared and a color fundus camera, as well as an automatic real-time tracking system that allows for a full automatic retinal fixation and threshold determination as well as automatic follow-up and differential maps determination, independently from fixation characteristics. The main technical characteristics of this instrument have been previously described in detail (Vujosevic et al., 2006; Midena et al., 2004; Midena et al., 2007). Roschneider et al. compared MP-1 and SLO microperimeters and found that both instruments analyzed retinal sensitivity and fixation characteristics, and the results obtained from both instruments were directly comparable. However, MP-1 is superior to SLO due to the automatic real-time alignment system, a larger field of (fundus) view ( $44^\circ \times 36^\circ$  MP1 versus  $33^\circ \times 21^\circ$  SLO) and color image (Rohrschneider et al., 2005).

The most relevant characteristics of advanced microperimetry performed with the MP-1 microperimeter may be briefly summarized as follows:

- Exact fundus-related stimulation
- Automatic eye-tracking system
- Automatic static and kinetic stimulation (with standardized or customized grids and centration)
- Normative age-related database (Midena et al., 2010)
- Age-related differential maps (local defect determination, shallow defects determination, etc. . .)
- Automatic follow-up and differential maps

Download English Version:

<https://daneshyari.com/en/article/2704392>

Download Persian Version:

<https://daneshyari.com/article/2704392>

[Daneshyari.com](https://daneshyari.com)