#### **CASE REPORT**

## Inferior Mesenteric Artery Side Branch for Selected Patients with Endovascular Aortic Aneurysm Repair

K. Pfister, P.M. Kasprzak \*, H. Apfelbeck, R. Kopp, M. Janotta, W. Schierling

Division of Vascular Surgery, University Medical Center Regensburg, Franz-Josef-Strauss-Allee 11, D-93053 Regensburg, Germany

Objective/Background: To report on our experience of the treatment of aortic aneurysms by custom-made, branched stent-grafts with an additional inferior mesenteric artery (IMA) side branch to preserve IMA perfusion in patients at risk for colon ischemia.

Methods: Three male patients (mean age 60 years) with a thoracoabdominal, pararenal, and infrarenal aortic aneurysm (AA), respectively, were treated by endovascular aneurysm exclusion using custom-made, branched stent-grafts with a side branch to the IMA for prevention of colon ischemia. Indications for selective IMA side branch perfusion were occlusions or high-grade stenosis of the visceral or hypogastric arteries.

Results: No colon ischemia and no neurological deficit were observed. All three IMA side branches were perfused and patent, as documented by computed tomography scan and duplex ultrasound postoperatively and after 12 months. Patency after 24 months was documented as 2/3.

**Conclusion:** Custom-made, branched stent-grafts are an endovascular option to preserve the IMA perfusion in selected, electively treated patients with an increased risk for insufficient colon perfusion due to stenosis or occlusions of visceral or hypogastric arteries.

© 2016 The Authors. Published by Elsevier Ltd on behalf of European Society for Vascular Surgery. This is an open access article under the CC BY-NC-ND license (http://creativecommons.org/licenses/by-nc-nd/4.0/).

Article history: Received 29 November 2015, Revised 18 January 2016, Accepted 20 February 2016,

**Keywords:** Branched stent-graft, Colon ischemia, Custom-made device, Endovascular aneurysm repair, Inferior mesenteric artery

#### **INTRODUCTION**

Colon ischemia is a well-known, major complication of open and endovascular aortic aneurysm repair (EVAR). The incidence is described as 1—6% in elective cases. Revascularization of the inferior mesenteric artery (IMA) is reserved for anatomical disorders such as superior mesenteric artery (SMA) and/or celiac trunk high-grade stenosis/occlusion or iliac/hypogastric artery occlusion. The IMA has an important collateral function for intestinal perfusion in these cases.

Donas et al. recently published the chimney technique as an endovascular tool to preserve IMA perfusion in patients with bilateral hypogastric artery occlusion.<sup>2</sup> In the associated commentary, Rancic stresses the importance of the correct patient selection and discusses the clinical relevance of an IMA revascularization because of the usually very functional collateral system.<sup>3</sup>

In the following case series, we report our experience on preservation of the IMA perfusion with custom-made, branched stent-grafts in selected patient with suspected compromised intestinal perfusion.

E-mail address: gefaess.chirurgie@ukr.de (P.M. Kasprzak).

2405-6553/© 2016 The Authors. Published by Elsevier Ltd on behalf of European Society for Vascular Surgery. This is an open access article under the CC BY-NC-ND license (http://creativecommons.org/licenses/by-nc-nd/4.0/). http://dx.doi.org/10.1016/j.ejvssr.2016.02.004

#### **CASE REPORT**

At our hospital, an EVAR procedure with an additional custom-made side branch to the IMA (Cook Medical, Europe Ltd., Limerick, Ireland) was performed in three patients with thoracoabdominal, pararenal, and infrarenal aortic aneurysms (AA), respectively. The idea to create an additional side branch to the IMA was initially developed because of a patient with an occlusion of the celiac trunk, a large IMA, and the need for a multibranched stent-graft (Kasprzak PM, personal communication). In all three patients, a 6-mm, downward-directed branch was used with extension to the IMA using 6 mm Fluency Plus stent-grafts (C.R. Bard GmbH, Karlsruhe, Germany), endolined with self-expandable uncovered stent. Diameter of the IMA was at least 4 mm in all patients.

A transfemoral and transaxillary approach was used to have a stabilizing wire from the upper limb to the femoral access. An F12 sheath (45 cm) was applied by the axillary access for catheterism of the IMA gate by a vertebral angiographic catheter. After catheterism, the Fluency Plus stent-graft was delivered via a stiff wire (Rosen Wire Guide [Cook Medical] or Amplatz Super Stiff Guidewire (Boston Scientific, Marlborough, MA, USA)]. Time for completion of the side branch to the IMA was about 20—30 minutes with an additional 15—20 mL contrast. Follow-up was done after 6 and 12 months, and annually thereafter by computed tomography (CT) scan and duplex ultrasound.

<sup>\*</sup> Corresponding author.

The retrospective data evaluation complies with the principles outlined in the Declaration of Helsinki and subjects gave informed consent (ethics committee approval REC number 12-101-0121). Table 1 summarizes patients' characteristics and treatment concepts.

Patient 1 (aged 64 years) had a thoracoabdominal AA of 6.6 cm with an occlusion of the celiac trunk, a high-grade stenosis of the SMA, stenosis of both renal arteries, and a large IMA of 4 mm. Distal aortic diameter was too small for a bifurcated graft (Fig. 1A). To preserve two visceral arteries for intestinal perfusion, procedure was planned as a left-sided, monoiliac, fourfold branched stent-graft with side branches to the SMA, both renal arteries, and the IMA (Fig. 1B). The patient also received a right-sided iliac plug and an iliofemoral left-to-right crossover bypass (Fig. 1B). The patient died 60 months after the procedure (not as a result of an aneurysm-related cause) with open branches to both visceral and renal arteries.

Patient 2 (aged 57 years) had a pararenal AA of 8.5 cm with a high-grade stenosis of the right hypogastric artery, an occlusion of the left iliac and hypogastric artery, and a large IMA of 4.5 mm (Fig. 2A). The procedure was initially planned as a right-sided, monoiliac, fivefold branched stent-graft with branches to the celiac trunk, SMA, renal arteries, the IMA, and a right-sided iliac branch device. Because of the high-grade stenosis of the right hypogastric artery and difficult anatomy, catheterism of this hypogastric artery failed (as previously feared) and the ostium had to be

covered by a stent-graft, resulting in occlusion of both hypogastric arteries (Fig. 2B). The patient additionally received a femorofemoral right-to-left crossover bypass (Fig. 2B). Follow-up could be continued until 24 months, demonstrating a patent IMA side branch. Afterwards, the patient was lost to follow-up and died 48 months after EVAR, most likely from heart failure or acute pulmonary embolization.

Patient 3 (aged 60 years) had a 4.9-cm, fast-growing, infrarenal AA, and suffered from bilateral buttock claudication due to bilateral hypogastric artery occlusion. The IMA (4.0 mm) showed collaterals to the left hypogastric and sacral arteries, and a large lumbar artery (4.0 mm) supplied blood flow to the pelvis and the spinal collateral network (Fig. 3A). An endovascular procedure was performed with a bi-iliac stent-graft and two additional branches to the IMA and the dominant lumbar artery (twofold branched stent-graft; Fig. 3B). Postoperatively and after 12 months, CT scan and ultrasound examination showed regular results with a patent IMA branch (Figs. 3B and 4). However, routine control after 24 months demonstrated an occlusion of the IMA stent-graft without any signs of stent-graft kinking (Fig. 3C). Collateral function via the patent lumbar branch remained unchanged after 24 months and after 64 months. So far, this patient has not been suffering from any clinical symptoms.

#### **DISCUSSION**

The incidence of colon ischemia in AA repair is described as 1-6%, and mortality is high (53% within 1 month) -

Table 1. Patients' characteristics and treatment concepts.

	Patient 1	Patient 2	Patient 3
Age (years)	64	57	60
Gender	Male	Male	Male
Comorbidities	Coronary artery disease	Coronary artery disease, severely impaired left ventricular function, peripheral artery disease	Coronary artery disease
Aneurysm morphology	Thoracoabdominal	Pararenal	Infrarenal
Aneurysm diameter (cm)	6.6	8.5	4.9
Additional anatomic features	Occlusion of celiac trunk, high-grade stenosis of SMA, stenosis of both renal arteries, distal aortic diameter to small for bifurcated graft, large IMA (4.0 mm)	hypogastric artery, large	Bilateral hypogastric artery occlusion, large lumbar artery (4.0 mm) and IMA (4.0 mm) as collaterals
Device implanted	Monoiliac: Branches and extension for SMA, both renal arteries, IMA (4-fold branched)	Monoiliac: Branches and extension for celiac trunk, SMA, both renal arteries, IMA (5-fold branched)	Bi-iliac: Branches and extension for dominant lumbar artery and IMA ( <b>2-fold branched</b> )
Fluency stent-graft	6 × 60 mm	6 × 80 mm	$6 \times 60 \text{ mm}$ (lumbar artery and IMA)
Additional surgical procedures	Iliofemoral crossover bypass (left > right), right-sided iliac plug (16 mm)	Femorofemoral crossover bypass (right > left), implantation of right-sided iliac branch device failed	Ø
Duration of spinal drainage (days)	) 3	3	Ø
Spinal cord ischemia	Ø	Ø	Ø
Postoperative colon ischemia	Ø	Ø	Ø
IMA patency (months)	60	24	12

SMA: superior mesenteric artery, IMA: inferior mesenteric artery.

### Download English Version:

# https://daneshyari.com/en/article/2911450

Download Persian Version:

https://daneshyari.com/article/2911450

<u>Daneshyari.com</u>