




Enfermedades Infecciosas y Microbiología Clínica

www.elsevier.es/eimc



Brief report

Usefulness of Genotype MTBDRplus assay in acid-fast bacilli positive smear specimens in Almeria, Spain



Miguel José Martínez-Lirola^a, María José Muñoz-Dávila^{a,*}, Darío García-de Viedma^b, Teresa Cabezas Fernández^c, Pilar Luzón García^d, Indal-Tb Group

^a Unidad de Gestión Clínica de Biotecnología, Complejo Hospitalario Torrecárdenas, Almería, Spain

^b Servicio de Microbiología, Hospital General Universitario Gregorio Marañón, Madrid, Spain

^c Empresa Pública Hospital de Poniente, El Ejido, Almería, Spain

^d Servicio de Análisis Clínicos (Microbiología), Hospital La Inmaculada, Huércal-Overa, Almería, Spain

ARTICLE INFO

Article history:

Received 6 November 2013

Accepted 24 February 2014

Available online 5 May 2014

Keywords:

Genotype MTBDRplus
Molecular susceptibility
Comparison
Time reduction

ABSTRACT

Introduction: The urgent need for operational research evaluating test performance in a real-world setting has been highlighted. The purpose of this study was therefore to evaluate the performance of MTBDRplus assay.

Materials: According to the reference method, of the 155 clinical specimens with valid results, 147 were susceptible to rifampicin (RIF) and isoniazid (INH), with 4 being multi-drug resistant (MDR) and 4 with isolated resistance to isoniazid (INH).

Results: The results of the MTBDRplus assay were 100% concordant for the MDR and mono-resistant to INH specimens. However, the MTBDRplus assay showed a resistance pattern to RIF in one specimen which was classified as susceptible by the reference method. The majority of the specimens (118/75.6%) were also tested using the MTBDRplus method after culture on Lowenstein-Jensen media, showing 100% agreement with the results of the test directly from the specimens. An MTBDRplus test result was available within an average of 8 days.

Conclusions: Overall, MTBDR results showed excellent results when compared with the reference method and achieved a significant time-reduction.

© 2013 Elsevier España, S.L.U. and Sociedad Española de Enfermedades Infecciosas y Microbiología Clínica. All rights reserved.

Utilidad del ensayo Genotype MTBDRplus en muestras clínicas con baciloscopia positiva en el sudeste español

RESUMEN

Introducción: Es importante evaluar el desarrollo de los ensayos moleculares en la práctica diaria del laboratorio de microbiología.

Materiales: Se incluyeron 155 muestras clínicas con resultado válido. De acuerdo con el método de referencia, 147 fueron sensibles a INH y RIF, 4 MDR y 4 presentaron resistencia aislada a isoniazida.

Resultados: Los resultados del ensayo Genotype MTBDRplus fueron concordantes 100% en la muestras MDR y con resistencia aislada a isoniazida. Sin embargo, el ensayo Genotype MTBDRplus demostró un patrón de resistencia a RIF en una muestra que fue sensible por el método de referencia. En 118 muestras (75.6%) también se realizó el ensayo Genotype MTBDRplus sobre la cepa obtenida tras cultivo en medio Lowenstein-Jensen, mostrando un 100% de concordancia con los resultados obtenidos por el test directamente en muestra clínica. De media, los resultados del ensayo Genotype MTBDRplus estuvieron disponibles en 8 días.

Palabras clave:

Genotype MTBDRplus
Susceptibilidad molecular
Comparación
Reducción tiempo de respuesta

* Corresponding author.

E-mail address: mariajose.munoz5@um.es (M.J. Muñoz-Dávila).

Conclusiones: En conjunto, el ensayo GenoType MTBDRplus mostró resultados excelentes cuando se comparó con el sistema de referencia y consiguió una reducción significativa en el tiempo de emisión de resultados.

© 2013 Elsevier España, S.L.U. y Sociedad Española de Enfermedades Infecciosas y Microbiología Clínica. Todos los derechos reservados.

Introduction

The rapid detection of Multidrug resistant tuberculosis (MDR-TB) ensures treatment with appropriate drugs, thereby reducing morbidity, mortality, economic costs, the further transmission of infection and the emergence of extremely drug resistant strains.¹ For these reasons, it is important for clinicians to receive rapid susceptibility results from the laboratory. By using molecular biological methods, such as line probe assays, to detect mutations that are associated with drug resistance against rifampicin (RIF) and isoniazid (INH), preliminary susceptibility results can be reported within a few days after receiving the sample. While tests such as the GenoType MTBDRplus assay offer great promise to improving MDR-TB detection and care, the urgent need for operational research evaluating test performance in a real-world setting has been highlighted.²

Thus, although the GenoType[®] MTBDRplus version 1.0 assay has been studied by several authors, this publication is the first of its kind in Almería province, where a moderate incidence of TB has been described (average incidence rate: 26.6 cases/100,000 inhabitants).³ The purpose of this study was therefore to evaluate the performance of MTBDRplus assay for the rapid detection of *Mycobacterium tuberculosis* and mutations causing RIF and high or low level INH resistance in an attempt to compare with the results obtained by our reference susceptibility method. The assay was performed directly in smear-positive pulmonary specimens and also in some of the *M. tuberculosis* isolates grown from the subculture from liquid media to Lowenstein-Jensen with objective of comparison. The MTBDRplus assay is compared to drug susceptibility testing performed by automated liquid culture system (BACTEC MGIT 960, BD, Sparks, MD, USA).

Materials and methods

Clinical specimens

Two hundred and fifty-two respiratory samples (225 sputa, 22 bronchial aspirates, 1 bronchoalveolar lavage and 4 gastric aspirates) from 247 patients (for each patient, only one specimen with the highest number of AFB on stain was taken into account) were prospectively selected from the samples collected between December 2009 and December 2011 during the clinical routine at the Torrecárdenas Hospital (Almería, Spain). Before they were tested, all samples were first digested and decontaminated by using the Kubica N-acetyl-L-cysteine NaOH method.⁴ Later, auramine-rhodamine acid-fast staining was performed on the concentrated sediment as we have already described.⁵ During the study period, samples-selection criteria were AFB on stain but only diagnostic specimens were included. There are more samples tested than patients included because in case of a MTBDRplus invalid result with the first specimen, another one was tested. An aliquot of the decontaminated specimens was cultured on BacT/Alert liquid medium (BioMérieux, Marcy l'Etoile, France) and Lowenstein-Jensen solid agar. The identification of *M. tuberculosis* complex was confirmed in cultures by the AccuprobeMTC assay (Gen-Probe, Inc., San Diego, CA).

Drug susceptibility

Testing for susceptibility to INH, RIF, pyrazinamide, ethambutol, and streptomycin was performed with the automated BACTEC MGIT 960 by using critical concentrations of 0.1 µg/ml for INH and 2 µg/ml for RIF at a level III Regional Reference Laboratory (Hospital Costa del Sol, Marbella, Spain).

GenoType MTBDRplus assay

The GenoType MTBDRplus assay (Hain Lifescience, Nehren, Germany) was performed directly from the 252 decontaminated smear positive respiratory specimens, before the culture results were obtained. Identification culture results found MTC in 177 (70.2%) specimens. Cultured sample on Lowenstein-Jensen material was available in 118 samples out of 177 (66.7%), in which the molecular assay was also performed. All tests were performed according to the manufacturer's instruction.

Genotypic characterization

One of the specimens showing disagreement between the molecular and the reference method was sequenced at a Reference Laboratory (Hospital Gregorio Marañón, Madrid, Spain).

Data analysis

The turnaround time of results from the molecular and the reference methods was compared. The data were analyzed by using SPSS software (version 15.0). Comparisons were done by chi-square analysis. A *P* value of ≤0.05 was considered significant.

Results

MTBDRplus assay results for the clinical specimens

Valid results were obtained for one hundred and fifty-five (155/252, 61.5%) specimens. Among the 252 specimens, identification culture results found MTC in 177 (70.2%). In 75 specimens, culture was negative (43; 17%), contaminated (7; 2.7%) or a NTM was isolated (25; 9.9%). Among 177 molecular tests performed in specimens yielding MTC in culture, 12.4% (22/177) were not interpretable due to: absence of MTC specific band (7 paucibacillar specimens, 4%), lack of any locus control band (14 specimens, 8%) and absence of more than one *rpoB* wild type band (1 specimen, 0.4%). None of the negative, contaminated or NTM cultured specimens showed an interpretable band pattern on the molecular test as all of them lacked the specific MTC band.

Among the 155 clinical specimens with valid results, 147 were fully susceptible by the reference method, being 4 MDR and 4 mono-resistant to INH. The specificity of the assay for the detection of resistance to RIF, INH and MDR was 99.3%, 100% and 100% respectively. The sensitivity for the detection of resistance to RIF, INH and MDR was 99.3%, 100% and 100%, respectively. Thus, the MTBDRplus assay showed a RIF resistance pattern in one specimen which was classified as susceptible by the reference method. In this specimen,

Download English Version:

<https://daneshyari.com/en/article/3400846>

Download Persian Version:

<https://daneshyari.com/article/3400846>

[Daneshyari.com](https://daneshyari.com)