



Effect of three cueing devices for people with Parkinson's disease with gait initiation difficulties



Paula J. McCandless^a, Brenda J. Evans^a, Jessie Janssen^{a,*}, James Selfe^a, Andrew Churchill^b, Jim Richards^a

^aAllied Health Research Unit, University of Central Lancashire, Preston, UK

^bSchool of Education and Social Science, University of Central Lancashire, Preston, UK

ARTICLE INFO

Article history:

Received 19 January 2015

Received in revised form 24 June 2015

Accepted 4 November 2015

Keywords:

Parkinson's disease

Gait initiation

Freezing of gait

Cueing devices

Assistive devices

ABSTRACT

Background: Freezing of gait (FOG) remains one of the most common debilitating aspects of Parkinson's disease and has been linked to injuries, falls and reduced quality of life. Although commercially available portable cueing devices exist claiming to assist with overcoming freezing; their immediate effectiveness in overcoming gait initiation failure is currently unknown. This study investigated the effects of three different types of cueing device in people with Parkinson's disease who experience freezing.

Methods: Twenty participants with idiopathic Parkinson's disease who experienced freezing during gait but who were able to walk short distances indoors independently were recruited. At least three attempts at gait initiation were recorded using a 10 camera Qualisys motion analysis system and four force platforms. Test conditions were; Laser Cane, sound metronome, vibrating metronome, walking stick and no intervention.

Results: During testing 12 of the 20 participants had freezing episodes, from these participants 100 freezing and 91 non-freezing trials were recorded. Clear differences in the movement patterns were seen between freezing and non-freezing episodes. The Laser Cane was most effective cueing device at improving the forwards/backwards and side to side movement and had the least number of freezing episodes. The walking stick also showed significant improvements compared to the other conditions. The vibration metronome appeared to disrupt movement compared to the sound metronome at the same beat frequency.

Conclusion: This study identified differences in the movement patterns between freezing episodes and non-freezing episodes, and identified immediate improvements during gait initiation when using the Laser Cane over the other interventions.

© 2015 Published by Elsevier B.V.

1. Introduction

Freezing of gait (FOG) remains one of the most common debilitating aspects of Parkinson's disease. It has been linked to injuries and falls and is a main contributory factor in reducing quality of life [1–4]. FOG causes temporary cessation of effective stepping and a sensation of “feet being glued to the floor” [1,5,6] and occurs when people turn (63%), initiate walking (23%), walk through narrow spaces (12%) and reach destinations (9%) [7].

There are multiple factors that can induce and overcome components of FOG [8,9] with pharmacological and surgical

intervention often unable to ameliorate symptoms [10]. The European guidelines for Parkinson's disease strongly recommend using cues for the improvement of walking speed; however, they weakly recommend against cueing of gait for improvement of freezing of gait [11]. This can be due to the limited literature that is available on this topic and the variety of cues used to improve freezing of gait. Transverse lines (TL) on the floor have been shown to improve gait in people with Parkinson's disease [12–15], including an increase in stride length [12,14,15] and improvement in gait initiation [12,15]. Other external cues, such as somato-sensory, visual and auditory stimuli, have also been used with mixed results; however, these studies focussed mainly on steady state gait and not on overcoming gait initiation failure [16,17].

Gait initiation failure or “start hesitation” is a component of FOG which is described as a difficulty in initiating gait in the

* Corresponding author at: Allied Health Research Unit, School of Sport, Tourism and The Outdoors, University of Central Lancashire, Greenbank 161, Preston PR1 2HE, UK. Tel.: +44 177 289 4560.

E-mail address: jjanssen@uclan.ac.uk (J. Janssen).

Unified Parkinson's Disease Rating Scale (UPDRS) [18]. Gait initiation is normally a stereotypical and unconsidered transition from stance into walking [19,20]. Giladi et al. explored the presence of motor blocks in a sample of 990 people with Parkinson's disease; 318 were found to have FOG, 86% of these had blocks in initiation of gait [5].

Studies on gait initiation failure are few. Jiang and Norman investigated the effects of visual and auditory cues on gait initiation in people with Parkinson's disease [12]. They found differences in maximum horizontal force between people with Parkinson's disease who freeze and do not freeze and between the different cues. The auditory cues used were rhythmic sounds matched to the participant's average step time and the visual cues were high-contrast transverse lines on the floor adjusted for the participant's height and first step length, which although beneficial has a limited practical value outside of the laboratory setting [21]. Moreover, the auditory cues in this study did not produce a significant difference, when compared to the no cue condition. Unfortunately, the authors grouped the individuals with and without gait initiation difficulty together, when studying the effect of the different cues, therefore, diluting the effect and the potential clinical relevance of the findings.

Van Wegen et al. investigated the use of a rhythmic somato-sensory cueing device attached to the wrist on gait initiation in people with Parkinson's disease. This showed that participants were able to modify their stepping pattern. The authors suggested that such cues draw attention to the act of walking [22]. Dibble et al. [23] considered the effects of different sensory cueing methods on gait initiation in people having Parkinson's disease. The cueing methods were a single and repetitive auditory signal from an electronic metronome and an electrical stimulus from a neuromuscular stimulator. Dibble et al. found that both these sensory cueing modalities had a negative effect on displacement of the body and swing limb [23]. Cubo et al. examined the effects of a metronome in 12 patients with freezing when in their "on" state and reached similar conclusions; walking time increased when using the metronome [24].

To date, no study has compared three types of cueing device (somato-sensory, visual and auditory cues) and their immediate effects on gait initiation in individuals who suffer from FOG episodes. The aim of this study was to explore which of three cueing modalities was most effective in reducing the FOG frequency and to determine how these cueing modalities facilitate gait initiation performance.

2. Method

Twenty participants were recruited from local Parkinson's Disease Society groups (14 males and 6 females), mean age 68 years (range 49–84 years) and 11.5 years (range 1–23 years) since diagnosis. Inclusion criteria were idiopathic PD diagnosed by a neurologist; ability to walk indoors without physical assistance; a score of two (occasional freezing when walking) or three (frequent freezing when walking, occasionally falls from freezing) on item 14 (freezing when walking) of the Unified Parkinson's Disease Rating Scale (UPDRS) [18]; adequate hearing and vision to perceive sound and visual cues; and no acute condition likely to cause gait impairment. For participants with motor fluctuations, timing of their data collection was based on their response to UPDRS item 14. Where participants met the freezing criteria only during an "off" period they voluntarily delayed their medication and were tested during an expected "off" phase. Ethical approval was gained from Cumbria and Lancashire NHS Research Ethics Committee ref: 08/H1015/76. All participants gave written informed consent according to the declaration of Helsinki [25] before entering the study.

Gait initiation data were collected at 100 Hz using a 10 camera Qualisys motion analysis system (Qualisys Medical AB, Gothenburg, Sweden). The calibrated anatomical system technique (CAST) was used to place and determine the movement of segments [26]. Anatomical markers were placed on the lateral and medial malleoli, epicondyles of femur and humerus, the greater trochanter, anterior and posterior superior iliac spines, head of acromium, ulnar styloid process and medial head of radius. Tracking markers were placed on the head of 1st and 5th metatarsals, calcaneus, anterior aspect of talus, clusters of four markers were placed on the shanks, thighs, arms and forearms. Force data were collected using two 400 mm by 600 mm AMTI force plates (BP400600 Advanced Mechanical Technology, Inc., USA) at 200 Hz. The raw data were then exported to Visual 3-D (C-Motion, Inc., USA) for processing. The movement and force data were filtered using a fourth order low pass Butterworth filter with a cut off frequency of 6 and 25 Hz, respectively. An area of 10 m by 3 m was covered in a plain blue coloured carpet matching the laboratory floor which covered the force plates and surrounding area. No wires or camera tripods were immediately in front to limit visual sensory information.

Participants were tested under five randomly assigned conditions; no cue, walking stick, a visual cue which was a laser line projected on the floor from a walking stick (Laser Cane, U-Step), an auditory cue provided by a metronome (Peterson BodyBeat Pulsing Metronome), and a somato-sensory cue using the same metronome set to vibration mode. Participants chose which hand to use for the walking stick and Laser Cane. The metronome was clipped to a belt at the back of the participant, while in vibration mode the vibration device was placed anteriorly over the right side of the pelvis so it could be felt easily. The auditory and somato-sensory cues were set at 70 beats/min and participants were asked if they felt comfortable with this setting; two chose to reduce the speed of repetition, to 60 and 50 beats/min.

Participants were asked to rise from a chair, stand briefly with one foot on each force plate, and then begin walking in their own time. This was to ensure that the instructions themselves did not act as a cue. The start of the initiation was defined by the initial movement of the centre of mass (COM) and the centre of pressure (COP) i.e. when the participants started or tried to start moving. The termination of the episode was defined as the swing foot leaving one of the force plates with the threshold of force plates set to 10 N. Two experienced neuro-physiotherapists independently determined if a freezing trial occurred. A freezing episode was determined when a consensus was reached. Before the cued trials the participants were asked to use the cue at the start of the initiation and in whatever way they felt would be most helpful. Data collection began while the participant was sitting and continued until they had walked a distance of up to 3 m, or as far as they were able. Participants wore their usual footwear. No instructions were given with regard to which foot to step off with. Participants rested as required in between each trial. A physiotherapist followed the participants for monitoring loss of balance and to prevent falls.

Outcome measures collected were percentage of freezing episodes, first step length, second step length, forward COM velocity, sideways COM velocity, number of forward/backward sways and the number of sideways sways, forward COP velocity (m/s) side to side COP velocity (m/s). The total body COM was found from a weighted sum of the COM of every segment of the body modelled, this included; feet, shanks, thighs, pelvis, trunk, head, arms and forearms.

Statistical analysis of the data was performed using SPSS Version 21. The statistical analysis of the biomechanical outcomes measures was conducted using an independent *t*-test to determine the differences between all episodes of freezing and non-freezing, as all episodes were included, this precluded the use of a paired

Download English Version:

<https://daneshyari.com/en/article/4055528>

Download Persian Version:

<https://daneshyari.com/article/4055528>

[Daneshyari.com](https://daneshyari.com)