

Accepted Manuscript

Title: A New Threshold Selection Method for X-ray
Computed Tomography for Dimensional Metrology

Authors: Tomasz Kowaluk, Adam Wozniak

PII: S0141-6359(17)30172-1

DOI: <http://dx.doi.org/doi:10.1016/j.precisioneng.2017.06.020>

Reference: PRE 6609

To appear in: *Precision Engineering*

Received date: 23-3-2017

Revised date: 22-6-2017

Accepted date: 26-6-2017



Please cite this article as: Kowaluk Tomasz, Wozniak Adam. A New Threshold Selection Method for X-ray Computed Tomography for Dimensional Metrology. *Precision Engineering* <http://dx.doi.org/10.1016/j.precisioneng.2017.06.020>

This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

A New Threshold Selection Method for X-ray Computed Tomography for Dimensional Metrology

Tomasz Kowaluk, Adam Wozniak

*Warsaw University of Technology,
Institute of Metrology and Biomedical Engineering
8 Boboli St., 02-525 Warsaw, Poland, wozniaka@mchtr.pw.edu.pl*

Highlights

- A new method of air and material separation for X-ray tomography for industrial dimensional metrology application was proposed;
- Factors influencing on the accuracy of reconstruction has been included on a calibration process;
- It was verified on five different tomographs for industrial application;
- Significant improvement in the accuracy of measurement was observed.

Abstract: The article presents a new method of air and material separation for X-ray tomography for engineering applications. It consists of executing additional calibration measurements and determining certain correction for air-material threshold values on the greyscale of the volumetric file obtained as a result of a tomographic measurement. Usability and effectiveness of this method has been experimentally verified on five different tomographs for industrial application. The proposed method performs better than existing well-known, automated algorithm ISO50% when measuring the volume as well as linear dimensions of workpieces made of different materials. Experiments shown considerable improvements of measurements for the tested conditions.

Keywords: measurement, X-ray tomography, air-material threshold

1. Introduction

Industrial computed tomography, which originates from medical tomography, is becoming a more common measurement technique in engineering applications. By applying higher energies and higher intensity of X-ray radiation it is possible to inspect complex machine parts made of materials with X-rays attenuation coefficient higher than human tissues. Tomography for engineering applications enables measurements unattainable by other metrological devices, e.g. non-destructive testing of the objects made of different materials such as metals, ceramics, plastics [1]. Investigation of internal structures allows detecting discontinuities in the objects. They are in the form of inclusions of other materials or voids in the form of porosity [2, 3]. It is possible to perform measurements of geometrically complex parts, to compare the point cloud achieved from the measurements with a CAD model or to create one in the process of so-called

Download English Version:

<https://daneshyari.com/en/article/5019065>

Download Persian Version:

<https://daneshyari.com/article/5019065>

[Daneshyari.com](https://daneshyari.com)