

Available online at www.sciencedirect.com



Tetrahedron Letters 46 (2005) 5731-5734

Tetrahedron Letters

Immobilisation and assessment of aniline dyes for non-fluorescent pH sensing applications

Lu Shin Wong and Mark Bradley*

School of Chemistry, University of Southampton, Southampton SO17 1BJ, U.K.

Received 28 May 2005; accepted 14 June 2005

Available online 7 July 2005

Abstract—The attachment of two aniline-based chromophores, Disperse Black 3 and Pararosaniline, to beaded controlled pore glass (CPG) and their testing as fibre optic based pH sensors is described. Synthetic methods for their attachment at specified loadings to CPG were developed. The Disperse Black 3 sensor displayed a rapid response time and a dynamic range between pH 1.0 and 2.5, while the Pararosaniline-bearing sensor gave an extremely slow response time but a large sensing range from pH 1 to 11. © 2005 Elsevier Ltd. All rights reserved.

The development of solid-state spectrometric pH sensors is the subject of intense investigation in many laboratories. 1-4 In the case of UV-vis sensors, organic indicator dyes supported on an insoluble macromolecular matrix interfaced to a UV-vis spectrometer to allow colorimetric changes with pH to be determined. As many indicators are aniline derivatives the development of procedures that allow the covalent attachment of these moieties at specified loadings would be useful. Here, the attachment of Disperse Black 3 1 and Pararosaniline base 2 to macroporous controlled pore glass (CPG) beads is described, allowing their integration into a sensor assembly⁵ and their UV-vis spectrometric responses with pH changes to be investigated. The azo-dye 1 was compared to the structurally related para-Methyl Red,⁵ while the triphenylmethane dye 2 displayed two colorimetric changes over the pH range (at pH 1-3 and 11-14)⁶ and offered the opportunity to extend the dynamic range of dye-based sensors, hitherto a major limitation of these sensors, which usually operate only within ± 2 units of the p K_i of the dyes used.⁷

For the synthesis of the CPG-supported indicators, (Scheme 1) aminopropyl-CPG conjugated to two 6-aminohexanoic acid spacers⁵ **3** was coupled with a mixture of *N*-Fmoc- and Cbz-protected glycine at various ratios to achieve a series of CPG beads with specified loadings (Table 1). Selective Fmoc deprotection and coupling with adipoyl dichloride followed by quenching of the supported acyl chloride (to avoid intrasite cyclisation or crosslinking) provided the carboxy-functionalised beads. These were coupled to the dyes using PyBOP since this reagent would not undergo guanidine formation⁸ with the excess aniline used. Other attachment methods for anilines such as 2,4,6-trichlorotriazine^{9,10} or isocyanates¹¹ proved unsuccessful for **2**.

After coupling, the dyes continued to respond to pH variations (Fig. 1). For 4, a single transition was observed as expected while with 5, three transitions were seen, with decolourisation occurring at the extremes of the pH range.

Under a microscope, a single 'sensor bead' of appropriate size was inserted in the sensor assembly⁵ between the ends of the optical fibre and the UV-vis spectra, pH profile and response times were examined in a 1 M HCl/NaOH system. In both cases, beads with a loading of $5 \mu \text{mol/g}$ were found to be sufficient for sensing purposes. For 4, the absorbance at λ_{max} (546 nm), corresponding to the protonated indicator, was plotted against pH and the best-fit sigmoid curve gave a p K_i of 1.71 (Figs. 2 and 3), similar to that of the *para*-Methyl

Keywords: pH sensor; Pararosaniline; Disperse Black 3; UV-vis; Controlled pore glass; Solid phase.

^{*}Corresponding author at present address: School of Chemistry, University of Edinburgh, Edinburgh EH9 3JJ, United Kingdom. Tel.: +44 (0)131 650 4730; fax: +44 (0)131 650 6453; e-mail: mbradley@staffmail.ed.ac.uk

Scheme 1. Attachment of indicators to CPG. Reagents and conditions: (a) Fmoc/Cbz-Gly-OH, DIC, HOBt, DMF, 4 h; (b) 20% v/v piperidine, DMF, 30 min; (c) adipoyl dichloride, DIPEA, DCM, N₂, 3 h; then 50% v/v H₂O, THF; (d) 1 or 2, PyBOP, DIPEA, NMP, 16 h.

Table 1. Loading of CPG after Cbz- and Fmoc-Gly-OH co-acylation^a

Molar ratio of Cbz-:Fmoc-Gly-OH	Fmoc loading post reaction	
	μmol/g	% relative to control
99:1	0.7	0.9
97:3	2.8	3.7
94:6	5.1	6.8
90:10	16	21
75:25	25	33
50:50	56	75
0:100 (Control)	75	100

^a 10 mol equiv of acid mixture relative to CPG loading




Figure 1. Visually observed colour change for beads in aqueous buffer.

Red sensor, which had a λ_{max} of \sim 540 nm and a p K_i of 2.07⁵ and to a solution of 1, which had a p K_i of \sim 1.78.¹²

For Pararosaniline, the single peak in the spectra of **2** in aqueous solution, ⁶ was resolved into two peaks for the supported dye **5** (Fig. 4) and could be related to the microenvironment around the dye¹³ afforded by the glass support.




Figure 2. UV-vis absorbance spectra of 4.

Plots of absorbance intensities of the two major peaks in the UV-vis spectra (460 and 581 nm) and the depression between the peaks (512 nm) did not indicate any obvious pK_i transition points (Fig. 5). Nevertheless, the sensor had a large response range from pH 1 to 11 although there was a completely unresponsive region between pH 7 and 8 while at both extremes of the pH range the absorbance dropped as the dye became de-colourised presumably due to the addition of the counter ion to the chromophore⁶ (5a and 5d, Scheme 2).

The response time of sensor **4** around its p K_i was comparable to that of the previously cited CPG-based sensors,⁵ with a $t_{95\%}$ of 8.7 s and a k of 0.348 s⁻¹ (Fig. 6). The response speed measurements of **5** were more

Download English Version:

https://daneshyari.com/en/article/5287432

Download Persian Version:

https://daneshyari.com/article/5287432

<u>Daneshyari.com</u>