



Contents lists available at ScienceDirect

American Journal of Infection Control

journal homepage: www.ajicjournal.org

AJIC
American Journal of
Infection Control

Brief Report

The practice of disinfection of high-speed handpieces with 70% w/v alcohol: An evaluation.

Flávia Morais Gomes Pinto PhD, RN ^{a,*}, Camila Quartim de Moraes Bruna PhD, RN ^a,
Tamara Carolina Camargo PhD, RN ^a, Máira Marques PhD RN ^a, Cely Barretos Silva MD ^b,
Suzete M. Sasagawa MD ^b, Lycia Maria Jenné Mimica PhD ^b,
Kazuko Uchikawa Graziano PhD, RN ^a

^a Department of Medical-Surgical Nursing, School of Nursing, University of São Paulo, São Paulo, Brazil

^b Department of Hospital Infection Control of Irmandade Santa Casa de Misericórdia de São Paulo, São Paulo, Brazil

Key Words:

Dental high-speed equipment
Ethanol
Sterilization
Infection control
Dental infection
Nursing

A high-speed handpiece is used in several dental procedures and the official recommendation for safe decontamination consists of rinsing with water, washing with detergent, mechanical friction, and sterilization; however, many professionals only apply 70% w/v alcohol without also cleaning the tool between patients. We performed an analysis of high-speed handpieces reprocessed only with 70% w/v alcohol and found that the methods used in clinical practice are not safe.

© 2016 Association for Professionals in Infection Control and Epidemiology, Inc. Published by Elsevier Inc. All rights reserved.

The risk of infection due to noncompliance with decontamination procedures of devices used in the health field is broadly recognized. Dental tools are potential sources of microbial dissemination among patients. They host and carry a universe of microorganisms, which greatly justifies safe management processing.¹⁻⁵

A high-speed handpiece (HSH) is used in several dental procedures, such as dental prophylaxis, preparing cavities for restoration, odontosections, and osteotomies.^{6,7} This technologically innovative tool not only has undeniable benefits in its application, but also has challenges in its decontamination due to its complexity. The official recommendation by scientific societies for safe HSH decontamination is rinsing with water, washing with detergent, and applying mechanical friction, followed by autoclave sterilization.^{8,9}

Currently, many professionals only apply 70% w/v alcohol without any other cleaning of the tool between patients.^{6,7,10} Dentists claim practicality and insufficient inventory to justify their practice of such methods.^{6,7,10}

The use of 70% w/v alcohol in the disinfection process, as opposed to sterilization, is expected to eliminate only microorganisms with

lower lethal power, because there is no destruction of bacterial spores.^{2,11} Alcohol is widely used as an intermediate surface-level disinfectant with both low cost and low toxicity. It is considered a nonspecific antimicrobial due to the multiplicity of its mechanism with toxic effects on microbial cells.²

There is no reliable data linking dental infections to deficient reprocessing of HSHs. Given the small number of infection notifications¹² and the disagreement among official recommendations and the practices observed in reprocessing HSH, this study aims to settle this doubt.

METHODS

The samples were collected from a dental practice in which the decontamination of HSHs was the use of alcohol without cleaning previous. The dental practice under study performs an average of 100 procedures per day.

To analyze the microbial load present on the external surface of HSHs, a carrier was chosen and validated to carry the microorganisms. This process was conducted because HSHs cannot be directly immersed into culture medium or liquid to extract the rinsed product. A sterilized hydrophilic cotton gauze (7.5 × 7.5 cm, 8 layers) (Cremer SA, Blumenau – SC, Brasil) was used because it has greater carrying capacity when compared with cotton swabs.

Validation of the capacity for microbial release from the gauze was performed by means of agitation and sonication. The gauze was previously contaminated with 1 mL suspension of *Serratia marcescens* (ATCC 14756), 10⁶ CFU/mL, followed by colony count by means of

* Address correspondence to Flávia Morais Gomes Pinto, PhD, RN, Department of Medical-Surgical Nursing, School of Nursing, University of São Paulo, Av Angélica, n° 1968 conj 21, Higienópolis, São Paulo, CEP 01228-200, Brazil.

E-mail address: flaviamoraisgp@gmail.com (F.M.G. Pinto).

Portions of this work are extracted from a doctoral thesis presented to the Adult Health Nursing Program at the Nursing School at University of São Paulo, São Paulo, SP, Brazil.

Conflicts of interest: None to report.

serial dilution after 24 hours and 48 hours kept at $37^{\circ}\text{C} \pm 2^{\circ}\text{C}$ in a bacteriologic incubator.

For the analysis of extraction capacity of the microbial load presented in HSH using the gauze, triplicated HSH samples were contaminated with 1 mL suspension containing 10^6 CFU *S. marcescens* (ATCC 14756). One piece of gauze embedded in 5 mL saline solution at 0.9% (0.9% SS) was rubbed on the HSH body, deposited in 300 mL 0.9% SS, and later sonicated and agitated. A total of 2 mL was removed from the rinse for serial dilution and later colony count.

The amount of times the gauze should be rubbed to ensure the maximum carrying of microorganisms from the HSH surface after disinfection was also evaluated. The HSHs were then contaminated with the same *S. Marcescens* suspension and after, disinfection with 70% w/v alcohol was performed, using different contact times (30, 60, and 90 seconds). Three gauzes embedded in 0.9% SS were then consecutively rubbed on the HSHs and deposited in flasks containing 300 mL 0.9% SS. After agitation and sonication, aliquots were collected to perform serial dilution and, later, colony count.

Decontamination of the water system in dental units

To prevent HSH contamination with microorganisms originating from the water system, validation of the decontamination of the water system used for cooling HSHs was performed. This procedure was applied before each sample was collected.

To disinfect the reservoir and the hose connecting the HSH, the water reservoirs were substituted by others containing a hydrogen peroxide-based disinfectant—Dentosept P Solution from Sirona (Sirona, Bensheim, Germany)—diluted at the 1:100 mL rate, resulting in the final concentration of 0.014%.¹³ With the HSH connected to the hose, the switch of the water line was activated for 2 minutes, with continuous disinfectant flow.¹³

After the contact time of the disinfectant with the water line, the original reservoir was supplied with sterilized distilled water and attached to the equipment. The switch was activated for 2 more minutes for complete removal of the disinfectant and residue.

Data collection

A sample size of 100 units (experimental group) was established, ensuring a 95% confidence interval and 9.8% absolute precision in the worst case, which happens when estimating a contamination incidence of 50%. The calculations were performed using the sampling book package from R software (R Foundation for Statistical Computing, Vienna, Austria).

The experimental group analyzed the microbial load on the HSH surface, without previous treatment, after disinfection with 5 mL 70% w/v alcohol. A gauze embedded in 5 mL 0.9% SS was rubbed on the HSH external surface to carry possible microorganisms.

Half of the samples ($n = 50$) were analyzed by the membrane filtering indirect method (method I). The gauzes were placed inside a previously sterilized container with 300 mL 0.9% SS. The container was submitted to 3 sessions of 5-second ultrasound (Ultrasonic Rinser USC-2800, Enge Solutions; Enge Solutions and Fanem, São Paulo, Brasil) and an additional 10 minutes of orbital agitation (Kline model 255; Fanem) at 160 rotations per minute to obtain the maximum dislocation of the possible microbial load collected by the gauze. In a biological protection cabin, the rinse was filtered in 3 equal parts, through membranes with $0.45\ \mu\text{m}$ porosity (Sterifil Holder; Millipore Industria e Comercio LTDA, Barueri, São Paulo, Brazil) and they were sown in specific culture medium for recovering aerobic and anaerobic microorganisms (Blood agar, *Mitis salivaris* & Anaeroinso; BD Difco, Spark, MD).

The other 50 samples were analyzed by direct immersion of the gauze in the culture medium (method II). After the gauze was rubbed

on the HSH's external surface, it was placed in a test tube containing 40 mL Fluid Thioglycollate Medium (BD Difco). The tube containing the gauze was agitated for 30 seconds in a vortex agitator to provide microbial detachment and greater contact of the gauze with the culture medium.

Positive and negative control groups were also formed, with 10 samples each. The samples in the negative control group were collected after standard cleaning and vapor sterilization by means of an autoclave cycle at 134°C for 5 minutes. The procedure to recover the microbial load followed that used in method I.

The samples of the positive control group were collected after the dentist attended to a patient without any previous disinfection treatment. After using the HSH, the same technique used in method I (quantitative) was used for microbial recovery.

All samples were incubated in a bacteriologic incubator at $37^{\circ}\text{C} \pm 2^{\circ}\text{C}$, with daily reading of the samples. In the presence of microbial growth in culture media, the samples were forwarded to quantification and/or identification of genus and species of the microorganisms. This procedure was in accordance with the routine at the Laboratory of Microbiology of the Hospital Infection Service at Irmandade Santa Casa de Misericórdia de São Paulo.

RESULTS

In the validation study of the method used to release microorganisms from the gauze used as a carrier, the average microbial recovery was 4.09×10^5 CFU and the microbial load carried by the gauze was 2.28×10^5 CFU. The application of 70% w/v alcohol to an HSH for 90 seconds presented the best contact time of the germicide in the reduction of the initial load, requiring only 1 gauze for microbial carry. The validation studies performed showed satisfactory microbial release from the gauze, as well as its effectiveness at extracting the microorganisms present on an HSH.

The samples analyzed by method I (filtration by membrane) presented positive growth in 27 out of 50 samples (54%). The total positive growth presented by method II (direct sowing of the gauze) was in 12 out of 50 samples (24%). The microorganisms found in both methods are described in Table 1.

The samples in the positive control group presented satisfactory results regarding the presence of microorganisms on HSHs (Table 2).

All 10 samples in the negative control group presented an absence of microbial growth.

DISCUSSION

The results of the present investigation rebuke the practice of disinfecting HSHs by friction with 70% alcohol without previous cleaning. This statement is substantiated by the survival of microorganisms that do not correspond to the fungal and bactericide action expected from the alcohol in the condition of intermediary disinfection that should eliminate mycobacteria, vegetative bacteria, fungi, and viruses.¹⁴

The microbial load present in the mouths of healthy human individuals was high, ranging from 10^6 – 10^9 CFU/mL (or organisms per gram) in saliva, 10^5 CFU/mL in the gingival sulcus and 10^{11} organisms per gram on the dental plate.^{15,16} Considering this bioburden, the reduction obtained in the study conditions was impressive, reducing to 10^0 – 10^2 CFU, but this did not eliminate microorganisms supposedly susceptible to the disinfectant used.

The microorganisms recovered, although of low pathogenic potential, can behave as opportunists capable of attacking a host when environmental and immunological conditions are favorable to the microorganisms and causing infection.¹⁷ This is yet another aspect that reinforces condemnation of the practice observed.

Download English Version:

<https://daneshyari.com/en/article/5566466>

Download Persian Version:

<https://daneshyari.com/article/5566466>

[Daneshyari.com](https://daneshyari.com)