

Available online at www.sciencedirect.com

ScienceDirect

www.elsevier.com/locate/semvascsurg

Secondary interventions after endovascular aneurysm sac sealing: endoleak embolization and limb-related interventions

Seyed Ameli-Renani, MBBS, FRCR*, and R.A. Morgan, MBChB, MRCP, FRCR, EBIR

Department of Radiology, St George's Hospital, Blackshaw Road, London, SW17 0PZ, UK

ARTICLE INFO

ABSTRACT

The Nellix endovascular aneurysm sealing system is a novel alternative to conventional endovascular aneurysm repair for aortic aneurysm management using paired balloon expandable endografts supported by polymer-filled endobags to achieve sealing and anatomic fixation. Part of the promise of endovascular aneurysm sealing is increased resistance to lateral and longitudinal forces and thus a potential for reduced rates of device-related failures, particularly endoleaks. Initial efficacy data on this device are encouraging, but our knowledge of its associated complications and their management is limited. Reported adverse events include Type 1 and 2 endoleaks, graft stenosis and occlusion. The aim of this article is to review the early experience of endovascular aneurysm sealing focusing on the incidence, significance, and management of device-related complications.

© 2016 Elsevier Inc. All rights reserved.

1. Introduction

The Nellix device (Endologix Inc., Irvine, CA) is a novel endograft for abdominal aortic aneurysm repair based on the concept endovascular aneurysm sealing (EVAS) using a bag-filled polymer to obliterate the sac. Early efficacy data have been encouraging, but our knowledge of EVAS complications and long-term durability are limited. In this report, we review the known device-related complications associated with EVAS and propose appropriate management strategies.

2. Overview of complications after EVAS

Several randomized trials and systematic reviews have shown EVAR to be associated with reduced early mortality

and morbidity compared to open surgery, but with a higher re-intervention rate due to graft-related complications, including endoleak and stent migration, which, in turn, require regular and long-term surveillance at increased cost [1–4]. Endoleaks and the ensuing risk of sac pressurization and rupture have been the main contributory factors in this regard.

The Nellix sealing system has been well described previously [5]. In brief, it comprises two balloon expandable 10-mm chromium-cobalt stent grafts mounted on identical 17Fr catheter-based delivery systems, one for each side, which provide a flow lumen in parallel from the non-aneurysmal aorta proximally to the iliac artery distally. Each stent is surrounded by a polymer-filled endobag that conforms to the shape of the aneurysm. The endobags fill the aortic lumen, thereby eliminating the otherwise open space of the sac.

*Corresponding author.

E-mail address: seyed.renani@stgeorges.nhs.uk (S. Ameli-Renani).

Positional stability is therefore provided by anatomic fixation, as well as a proximal and distal seal. It is hoped by the manufacturers that this would minimize any potential longitudinal and lateral movements, which can contribute to Type 1 endoleaks [6]. In addition, filling of the sac by the endobags may tamponade any side branch flow that would produce any Type 2 endoleaks [5]. EVAS also enables treatment of a wider range of aortic anatomies than conventional EVAR [7].

Although this device has been used in a relatively large number of patients since its introduction in 2010, knowledge of the potential complications after Nellix is limited.

The first published data on the outcomes of Nellix in 2011 included a small series of 34 patients [8], which encompassed 21 cases analyzed in an earlier publication [9]. These cohorts included two cases of endoleaks, a transient Type 1a (EL1a) and a Type 1b (EL1b) treated by placement of an extension endograft. However, in the last 2 years, there have been five publications of single or multicenter outcomes [10–14]. The main device-specific complications within these reports are Type 1 (EL1) and Type 2 endoleaks (EL2), limb stenosis, and limb occlusion. These are summarized in Table 1 and will be reviewed in more detail.

2.1. Endoleaks

Endoleaks are categorized into five main types. Type 1 arising at the proximal (1a) or distal (1b) endograft attachment site; Type 2 due to filling of the sac from retrograde flow in the side branch vessels; Type 3 from fabric tear or modular disconnection; Type 4 is usually transient and seen at the time of implantation in anticoagulated patients due to graft porosity; and Type 5, defined as an increase in the sac size in the absence of a visible endoleak (endotension).

The unique design of the Nellix endograft relies on endobags for device stability rather than the radial force and barb engagement provided at the proximal and distal fixation sites in conventional EVAR. As will be described, the only reported

endoleaks with the Nellix device are Types 1 and 2. The absence of Type 3 endoleak is explained by its single-unit separate right and left stent design without modular components.

2.1.1. Type 1a endoleak

A review of the literature shows a Type 1A endoleak rate of 0% to 3% after Nellix implantation. In the largest cohort of patients to date, Böckler and colleagues [11] reported on their experience in 171 cases performed at multiple European centers. They found no intraoperative EL1a, but 5 EL1a on follow-up, three seen at 1 month, and two at 6 months. One of these resolved spontaneously, two were embolized and two were observed. There was no difference in the aortic neck length in cases with or without EL1a (22 ± 12 mm vs 28 ± 15 mm; $P = .39$). Four of five EL1a were seen in the 116 cases within the manufacturer's instructions for use (IFU) and 1 within the 55 cases outside the IFU.

Brownrigg et al [12] reported four EL1a among a cohort of 105 patients, which were all seen in patients with adverse proximal necks and all successfully embolized with Onyx (Covidien, Irvine, CA) and coils. The 30-day outcome results of the US pivotal trial involving 150 patients reported one EL1a treated by coil embolization [10]. Reijnen et al [14] reporting on the global Nellix experience of ruptured (28 cases) and symptomatic ($n = 30$) aneurysms found a case of EL1a in the ruptured group in conjunction with distal migration of the stents, and subsequent fatal second rupture. An additional case of EL1a in the symptomatic group was successfully embolized.

Finally, Zerwes and colleagues [13] reported one EL1a among a cohort of 50 patients, which was associated with a partial rupture of the aneurysm sac during device implantation. The endoleak was successfully eliminated by implanting two additional Nellix endografts in combination with chimney grafts into both renal arteries.

Table 1 – Published outcome data on endoleaks and limb occlusions after Nellix endovascular aneurysm sealing.

First author, year	Patient population, study type, and population cohort	Aortic morphology, n (%)	Type Ia	Type Ib	Type II	Limb stenosis/occlusion
Reijnen, 2016 [14]	n = 58; multicenter retrospective, symptomatic (n = 30) and ruptured (n = 28) aneurysms	30 (52) outside IFU	2	1 (bilateral)	0	0
Böckler, 2015 [11]	n = 171; 5 months median (0–14); multicenter retrospective, elective cases	55 (32) outside IFU	5	3 (1 bilateral)	4	5 limb stenoses and 8 limb occlusion
Brownrigg, 2015 [12]	n = 105; single-center prospective, elective cases	72 (69) with adverse proximal necks	4	0	0	3 limb stenoses
Zerwes, 2016 [13]	n = 50; single-center prospective, elective cases	14 (28) outside IFU	1	0	1	3 graft stenoses (2 associated with distal embolism)
Carpenter, 2016 [10]	n = 150; multicenter prospective, elective cases	All within IFU	1	0	8	0

IFU, instructions for use.

Download English Version:

<https://daneshyari.com/en/article/5621713>

Download Persian Version:

<https://daneshyari.com/article/5621713>

[Daneshyari.com](https://daneshyari.com)