



The Effect of Skin Pigmentation on the Accuracy of Pulse Oximetry in Infants with Hypoxemia

Elizabeth E. Foglia, MD MSCE^{1,2}, Robin K. Whyte, MB³, Aasma Chaudhary, BS, RRT², Antonio Mott, MD⁴, Jodi Chen, MD⁵, Kathleen J. Propert, ScD⁶, and Barbara Schmidt, MD, MSc^{1,2,6}

To compare pulse oximetry measurement bias between infants with hypoxemia with either dark skin or light skin with Masimo Radical 7 and Nellcor Oximax. There was no significant difference in systematic bias based on skin pigment for either oximeter. (*J Pediatr* 2017;182:375-7).

Pulse oximetry provides a safe and noninvasive method to monitor arterial oxygen saturation (SaO_2) levels continuously in children who are critically ill.¹ Studies of adults have demonstrated that pulse oximetry oxygen saturation (SpO_2) consistently overestimates measured SaO_2 in individuals with hypoxemia who have dark skin.^{2,3} The impact of skin pigmentation on the accuracy of pulse oximetry in infants with hypoxemia is unknown.

The primary study objective was to compare the bias in pulse oximetry measurement between infants with hypoxemia with either dark skin or light skin. The secondary objective was to determine the accuracy and precision of 2 commonly used oximeters from 2 different manufacturers in infants with hypoxemia.

Methods

This was a prospective, cross-sectional study conducted at the Children's Hospital of Philadelphia. Infants were enrolled with cyanotic congenital heart disease and baseline oxygen saturation $<90\%$ and an indwelling arterial line; we excluded infants with elevated methemoglobin levels, hypotension, or anemia. Each subject was only enrolled once. Written informed consent for study participation was obtained from eligible infants' parents. This study was conducted with approval from the institutional review board of Children's Hospital of Philadelphia.

Arterial co-oximetry samples were analyzed with the Siemens Rapidlab 1265 (Siemens Healthcare, Erlangen, Germany). The Rapidlab uses a polychromator to measure light absorbance at several wavelengths from human blood to detect and quantify total hemoglobin and its derivatives. According to the manufacturer, fetal hemoglobin (tested at 20%, 40%, and 85% levels) does not interfere with oxyhemoglobin or deoxyhemoglobin measurement results.⁴

Pulse oximetry was measured simultaneously with 2 oximeters. The Nellcor Oximax (Covidien, Boulder, Colorado) was used clinically in conjunction with the GE Solar 8000

patient monitor (Version 5.4, Neonatal ICU mode; GE Healthcare, Little Chalfont, United Kingdom) with an averaging time of 6 seconds. In addition, a study-specific Masimo Rainbow SET Radical 7, (V7.9.1.0, with LNOP sensor; Masimo Corp, Irvine, California), was programmed with an averaging time of 10 seconds.

The pulse oximeter sensors were placed on the infant's extremities under the observation of operational rules to prevent a differential from right to left or bidirectional shunting across the ductus arteriosus. An opaque blanket was placed over the pulse oximeter sensors, and a study team member confirmed pulse oximetry measurement stability and signal quality. The study team member recorded the displayed pulse oximetry measurement at the same time that an arterial blood sample was obtained from an indwelling arterial catheter. The blood sample was sent immediately to the hospital laboratory for co-oximetry measurement by technicians who were not connected with the study team.

Baseline demographic data and clinical variables that could affect pulse oximetry accuracy were extracted from the medical record. Consistent with previous studies,^{5,6} skin pigment group was allocated with the Munsell System Soil Color Chart (2009 Revision, Munsell Color, Grand Rapids, Michigan), Hue 7.5YR (Figure 1; available at www.jpeds.com). The value score that best matched the subject's skin on the dorsal surface of the distal extremities was recorded.

As described previously,⁷ the accuracy and precision of the pulse oximeters were assessed with the following terms. Bias was defined as $\text{SpO}_2 - \text{SaO}_2$ for a given pair of measurements in a subject. Precision was defined as 1 SD of the mean distance between individual data points (plotted SpO_2 vs SaO_2) and the best-fit line. Accuracy root mean square (Arms)

Arms	Accuracy root mean square
SaO_2	Arterial oxygen saturation
SpO_2	Pulse oximetry oxygen saturation

From the ¹Department of Pediatrics, Perelman School of Medicine, University of Pennsylvania, Philadelphia, PA; ²Division of Neonatology, Children's Hospital of Philadelphia, Philadelphia, PA; ³Department of Pediatrics, Dalhousie University, Halifax, Nova Scotia, Canada; ⁴Department of Pediatrics, Baylor Medical College, Houston, TX; ⁵Department of Anesthesiology and Critical Care Medicine, Children's Hospital of Philadelphia, Philadelphia, PA; and ⁶Center for Clinical Epidemiology and Biostatistics, Perelman School of Medicine, University of Pennsylvania, Philadelphia, PA

Supported by Thrasher Research Fund (Early Career Award [9222; to E.F.]) and the Eunice Kennedy Shriver National Institute of Child Health and Human Development (1T32HD060550-01A1 [to E.F.]), and National Institutes of Health Loan Repayment Program (to E.F.). The authors declare no conflicts of interest.

0022-3476/\$ - see front matter. © 2016 Elsevier Inc. All rights reserved.
<http://dx.doi.org/10.1016/j.jpeds.2016.11.043>

accounts for both the accuracy and precision of pulse oximetry measurement. Arms is calculated as follows:

$$\text{Arms} = \sqrt{\frac{\sum_{i=1}^N (\text{SpO}_{2i} - \text{SaO}_{2i})^2}{n}}$$

The US Food and Drug Administration requires that pulse oximeters have Arms $\leq 3\%$ in SpO₂ range from 70% to 100%,⁸ in conformance with the International Organization of Standardization. Arms were calculated for all measurements and for measurements when SaO₂ $\geq 70\%$.

Using Stata 14.0 (StataCorp, College Station, Texas), we assessed differences between groups with light and dark pigment for continuous variables using *t* test (parametric) and the Wilcoxon rank sum test (nonparametric). For dichotomous variables, the χ^2 test and Fisher exact test were used as appropriate. Differences in bias were compared between groups with light and dark pigment (Student *t* test) and between oximeters (paired *t* test). Linear regression was used to control for baseline characteristics that significantly differed between pigment groups in univariable analysis with *P* value $< .05$. The study had 80% power to detect 1 SD difference in measurement bias between pigment groups.

Results

Between April 25, 2013, and December 23, 2015, 101 infants were screened, and 36 were enrolled (Figure 2; available at www.jpeds.com). There were 21 infants in the group with light pigment and 14 infants in the group with dark pigment. One

infant was classified as intermediate pigment (these data were not included in the skin pigment analysis but were included in the assessment of accuracy). Two Nellcor measurements from the group with light pigment were excluded (one because of concern for right-to-left shunting across the ductus arteriosus and one for poor signal quality).

The group with dark pigment was significantly older at the time of the procedure and had a greater mean arterial blood pressure. Infants were not hypothermic, hypotensive, acidotic, or anemic at the time of the procedure (Table). Measured SaO₂ values ranged from 60% to 92%. Measurement bias did not differ significantly between groups with light and dark pigment for either oximeter, with or without adjustment for age at the time of the procedure and mean arterial blood pressure (Table).

The mean Masimo measurement bias was 0.8% (SD 4.2%). The precision of Masimo measurement was 2.4%, and calculated Arms was 4.2%. For 32 measurements with SaO₂ values $\geq 70\%$, the calculated Arms was 4.2%. The mean measurement bias for the Nellcor was 3.9% (SD 5.0%), and the precision (SD) was 2.0%. The calculated Nellcor Arms for all subjects was 6.3%. For 30 measurements with SaO₂ values $\geq 70\%$, the calculated Nellcor Arms was 5.4%. Mean Nellcor bias was significantly greater than Masimo bias, *P* = .0005 (Figure 3; available at www.jpeds.com).

Discussion

We undertook this study to evaluate pulse oximetry performance under 2 conditions: hypoxemia and darkly pigmented skin. We did not find evidence of systematic bias in pulse

Table. Baseline characteristics and measurement bias according to skin pigment

	Light pigment (n = 21)	Dark pigment (n = 14)	<i>P</i> value
Demographic characteristics			
Gestational ages, wk, median (IQR)	39 (38, 40)	38 (37, 39)	.19
Ages at procedure, d, median (IQR)	6 (5, 66)	118 (80, 141)	<.001
Males	14 (67%)	7 (50%)	.32
Exposure to exogenous red blood cells	16 (76%)	14 (100%)	.07
Hemoglobinopathy	0	1 (7%): Hemoglobin FAS	.40
Diagnoses			.99
Hypoplastic left heart	7	6	
Tetralogy of Fallot	3	1	
Pulmonary atresia	4	2	
Tricuspid atresia	1	1	
Truncus arteriosus	1	1	
Transposition of the great arteries	1	1	
Double outlet right ventricle	1	1	
Pulmonic stenosis	1	1	
Other	2	0	
Clinical characteristics at procedure			
Hemoglobin, g/dL, mean (SD)	14.5 (1.4)	13.8 (1.9)	.23
pH, mean (SD)	7.4 (0.1)	7.4 (0.1)	.73
Skin temperature, °C, mean (SD)	37.0 (0.5)	36.9 (0.6)	.49
Mean arterial blood pressure, mm Hg, mean (SD)	55 (14)	67 (12)	.01
Milrinone	10 (48%)	7 (50%)	.89
Dopamine	2 (10%)	0	.51
Measurement biases, mean (SD)			
Masimo Radical 7	0.2% (3.8%) (n = 21)	1.6% (4.8%) (n = 14)	.39*
Nellcor Oximax	3.0% (5.0%) (n = 19)	5.4% (5.1%) (n = 14)	.39*

*Adjusted in linear regression for postnatal age at procedure and mean arterial blood pressure.

Download English Version:

<https://daneshyari.com/en/article/5719701>

Download Persian Version:

<https://daneshyari.com/article/5719701>

[Daneshyari.com](https://daneshyari.com)