## Accepted Manuscript

Title: Development and validation of an ELISA kit (YF MAC-HD) to detect IgM to yellow fever virus

Author: Alison Jane Basile Christin Goodman Kalanthe Horiuchi Janeen Laven Amanda J. Panella Olga Kosoy Robert

S. Lanciotti Barbara W. Johnson

PII: S0166-0934(15)00305-5

DOI: http://dx.doi.org/doi:10.1016/j.jviromet.2015.08.025

Reference: VIRMET 12879

To appear in: Journal of Virological Methods

Received date: 7-7-2015 Revised date: 31-8-2015 Accepted date: 31-8-2015

Please cite this article as: Basile, A.J., Goodman, C., Horiuchi, K., Laven, J., Panella, A.J., Kosoy, O., Lanciotti, R.S., Johnson, B.W., Development and validation of an ELISA kit (YF MAC-HD) to detect IgM to yellow fever virus, *Journal of Virological Methods* (2015), http://dx.doi.org/10.1016/j.jviromet.2015.08.025

This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.



### ACCEPTED MANUSCRIPT

#### Highlights

- A yellow fever IgM ELISA assay kit (YF MAC-HD) based upon the CDC MAC-ELISA was developed
- Assay time is reduced from 2 days to 3.5 h
- The kit is >90% accurate when compared to the CDC MAC-ELISA
- Kits are stored at 4 °C and can tolerate temporary suboptimal storage or shipping conditions
- The kit will reduce materials wastage and improve quality of YFV serologyin endemic areas

#### Download English Version:

# https://daneshyari.com/en/article/6133042

Download Persian Version:

https://daneshyari.com/article/6133042

<u>Daneshyari.com</u>