



Long-distance walking effects on trans-tibial amputees compensatory gait patterns and implications on prosthetic designs and training

L.F. Yeung, Aaron K.L. Leung, Ming Zhang, Winson C.C. Lee *

Department of Health Technology and Informatics, The Hong Kong Polytechnic University, Hung Hom, Kowloon, Hong Kong

ARTICLE INFO

Article history:

Received 27 July 2011

Received in revised form 4 October 2011

Accepted 9 October 2011

Keywords:

Gait
Compensatory mechanism
Long-distance walking
Trans-tibial amputees

ABSTRACT

Trans-tibial amputees are advised to walk as much as able people to achieve healthy and independent life. However, they usually have difficulties in doing so. Previous researches only included data from a few steps when studying the gait of amputees. Walking over a long distance was rarely examined. The objective of this study was to investigate the changes in spatial-temporal, kinetic and kinematic gait parameters of trans-tibial amputees after long-distance walking. Six male unilateral trans-tibial amputees performed two sessions of 30-min walking on a level treadmill at their self-selected comfortable speed. Gait analysis was undertaken over-ground: (1) before walking, (2) after the 1st walking session and (3) after the 2nd walking session. After the long-distance walking, changes in spatial-temporal gait parameters were small and insignificant. However, the sound side ankle rocker progression and push-off were significantly reduced. This was due to the fatigue of the sound side plantar flexors and was compensated by the greater effort in the prosthetic side. The prosthetic side knee joint showed significantly increased flexion and moment during loading response to facilitate the anterior rotation of the prosthetic shank. The prosthetic side hip extensors also provided more power at terminal stance to facilitate propulsion. Endurance training of the sound side plantar flexors, and improvements in the prosthetic design to assist anterior rotation of the prosthetic shank should improve long-distance walking in trans-tibial amputees.

© 2011 Elsevier B.V. All rights reserved.

1. Background

The ultimate goal of rehabilitation programs for trans-tibial amputees is to facilitate the best prosthetic function which allows them to undertake daily activities and achieve the highest possible activity level. However, amputees cannot be as physically active as able people [1], and are more susceptible to cardiovascular disease, including hypertension, hyperlipidemia and myocardial infarction [2]. The World Health Organization (WHO) states that people with disability are advised to undertake regular exercise and walking as much as able people [3]. Daily walking for over 30 min is recommended in order to attain a healthy life [4].

It has been suggested that in the modern community amputees are required to walk for longer distance and at higher velocity to achieve independent life [4]. However, another survey revealed that among 500 amputees interviewed 84% would walk outdoors for only 1–2 km at a time, and 60% of amputees could not tolerate walking for a long distance because they felt tired easily [5]. There is a need for better understanding of amputee gait over longer distance walking.

Previous gait analysis studies reported that trans-tibial amputees walk with shorter sound side step length due to the lack of prosthetic side ankle joint motion and greater ground reaction forces (GRF) at the sound side because of the sensation of pain at the prosthetic side [6–8]. As a result of perceived instability, they have decreased range of motion at the prosthetic side knee joint during walking [6–8]. This asymmetric gait pattern could lead to musculoskeletal disorders such as knee osteoarthritis at the sound side and scoliosis [13–15]. They walk more slowly but expend greater energy than able people when walking at self-selected comfortable speeds [9–12]. A variety of gait conditions have been investigated [16]: different gait initiation strategies, walking speeds, walking surfaces, alignments, types of amputation, prosthetic knee and feet designs as well as comparisons between fallers and non-fallers. All these studies provided valuable insights on how the amputee gait could be improved. Yet most of the published researches examined amputee gait for a few steps only. Gait changes when walking for longer distances have not been investigated.

The objective of this study was to investigate the changes in spatial-temporal, kinetic and kinematic gait parameters of trans-tibial amputees after performing long-distance walking. Such investigation could explain why amputees cannot walk for a long distance. It could provide useful information for developing

* Corresponding author. Tel.: +852 2766 4942; fax: +852 2362 4365.
E-mail address: htwinson@inet.polyu.edu.hk (Winson C.C. Lee).

strategies suitable for amputee training and for prosthetic design improvements to facilitate longer distance walking.

2. Methods

2.1. Subjects

Six male subjects between 39 and 62 years of age, who had unilateral trans-tibial amputation performed at a minimum of 8 years prior to recruitment participated in the study. Their mean height was 170 cm (SD = 3.4) and their mean body weight was 75 kg (SD = 4.7). Immature residual limb with large volume change and walking tolerance of less than an hour were exclusion criteria. Detailed characteristics of the subjects are shown in Table 1. They were recruited from the Jockey Club Rehabilitation Engineering Clinic. Ethical approval was obtained from the Human Subject Ethic Sub-committee of The Hong Kong Polytechnic University and all subjects provided informed consent to participate.

2.2. Testing protocol

Subjects performed two consecutive sessions of 30-min walking at their self-selected comfortable speed on a level treadmill. The comfortable speed was achieved by instructing the amputees to adjust to the appropriate speed on the treadmill. Subjects were restricted from holding the handrail of the treadmill. For safety reasons, the test was monitored by a professional prosthetist and subjects could stop the test at any time. Collection of gait data when the subjects walked over-ground was undertaken (1) before the walking (baseline), (2) after the 1st 30-min walking and (3) after the 2nd 30-min walking. Trials were repeated until at least five samples of successful walking steps were collected, in which the whole foot completely landed within the force plate. All successful trials were used in the analysis. After treadmill walking, 2–3 min of time was given to the subjects to walk over-ground to get familiar with the changing environment. Each gait data collection session was completed within 5 min. Amputee subjects reported the change in lower-limb fatigue level and residual limb pain levels after 1-h walking using the visual analogue scale (VAS) marked on a 10-cm line [17].

2.3. Apparatus

Subjects used their own prostheses, with patellar-tendon-bearing sockets, throughout the experiment (Table 1). An eight-camera motion capture system (Vicon, Oxford Metrics) was used for three-dimensional motion analysis. Infra-red reflective markers were placed bilaterally on the dorsum of the foot, heel, lateral and medial malleoli, lateral and medial femoral condyles, greater trochanter, anterior superior iliac spine, posterior superior iliac spine, iliac crest and mid-shank and thigh. GRF was measured using two force plates (Advanced Mechanical Technology, Inc.) embedded midway on a straight 8-m walkway. Kinematic and GRF data were collected at 200 and 1000 Hz, respectively.

2.4. Data analysis

Gait data were analyzed using Visual3D (C-motion, Inc.). All kinematic and kinetic data were low-pass filtered using a 4th-order Butterworth filter with a cut-off frequency of 6 Hz. GRF in the anterior direction (propulsive force), in the posterior direction (braking force), in the medial–lateral direction (GRFy), in the vertical direction (GRFz), amplitudes of center of body mass (CoM) vertical movements, angle, moment and power at the ankle, knee and hip joints, walking speed, step lengths, stance time and cadence were measured. GRFz ratio, propulsive force ratio and braking force ratio were defined as the peaks of GRFz and GRFy in the prosthetic side divided by those of the sound side.

Kinetic data were normalized to the body mass. Local maxima and points of interest in kinetic and kinematic data were determined within the gait cycle for each successful walking trial. All spatial–temporal, kinetic and kinematic quantities were averaged in each test session, and then averaged across subjects. The Bonferroni-adjusted *t*-test for dependent sample was used to determine if there were significant differences between (1) the baseline and after the 1st walking

session, as well as (2) the baseline and after the 2nd walking session. Significance level was set at $p < 0.05$.

3. Results

Before the two walking sessions, the peak sound side GRFz, propulsive force and braking force were significantly greater than that of the prosthetic side, with GRFz ratio of 0.87 ($p = 0.016$), propulsive force ratio of 0.49 ($p = 0.003$) and braking force ratio of 0.64 ($p = 0.009$), respectively. Stance time of the sound side was significantly longer than the prosthetic side by 6% ($p = 0.023$). Step length of the prosthetic side was significantly longer than the sound side by 13% ($p < 0.001$).

All six amputees were able to complete the two sessions of 30-min walking. After each walking session, changes in amplitudes of CoM and spatial–temporal gait parameters including walking speed, cadence, sound side and prosthetic side step lengths and stance time were very small and insignificant compared to the baseline (Table 2). The fatigue score in VAS for trans-tibial amputees was 7.9 ± 0.8 after the 2nd 30-min walking, which was significantly higher ($p < 0.001$) than the baseline of 3.0 ± 0.8 (less fatigue). There were no significant changes in the residual limb pain score.

There were significant reductions in sound side peak propulsive force by 15% (1st session: $p = 0.024$; 2nd session: $p = 0.022$) and peak braking force by 9% (1st session: $p = 0.017$; 2nd session: $p = 0.005$) (Fig. 1). The prosthetic side had significantly greater peak propulsive force by 21% ($p = 0.02$) while it had significantly less peak braking force by 19% ($p = 0.006$) after the 2nd 30-min walking (Fig. 2). After the two walking sessions, the propulsive force ratio increased significantly by 39% ($p = 0.003$).

After the 1st and 2nd 30-min walking, the prosthetic side knee joint had more flexion from 0% to 60% of the gait cycle. The first peak flexion angle was one-folded significantly higher ($p < 0.025$) than the baseline after the 2nd 30-min walking. No significant difference was found in sound side knee joint angles. However, from 10% to 30% of the gait cycle the sound side knee joint moment decreased with the peak value significantly reduced by 26% ($p = 0.025$) after the 2nd 30-min walking. During the same period of gait cycle, the prosthetic side knee joint moment increased significantly by 42% after the 1st 30-min walking ($p < 0.025$) and 67% after the 2nd 30-min walking ($p = 0.02$). At 60% of the gait cycle, the sound side knee joint moment decreased significantly by 15% ($p = 0.005$).

After the 2nd 30-min walking, there was significantly less sound side dorsiflexion at around 30% of the gait cycle ($p < 0.025$). There were also a trend toward 4% less dorsiflexion at around 50% of the gait cycle and 8% less plantar flexion at around 70% of the gait cycle. There was a reduction in sound side ankle joint moment from 25% to 62% of the gait cycle with the peak moment significantly reduced by 14% ($p = 0.025$).

The sound side ankle power absorption at 20% of the gait cycle decreased significantly ($p < 0.025$) after the 2nd 30-min walking.

Table 1
Characteristics of participated amputee subjects.

Amputee participants	Gender	Age (years)	Height (cm)	Body mass (kg)	Time since amputation (years)	Amputated limb	Type of prosthetic foot	Reason for amputation
A	M	61	168	66	39	L	SACH	Traumatic
B	M	47	170	75	9	R	SACH	Traumatic
C	M	54	167	79	45	L	SACH	Traumatic
D	M	56	168	78	36	R	ESAR	Traumatic
E	M	62	176	75	22	R	SACH	Traumatic
F	M	39	167	76	17	R	SACH	Vascular
Mean (SD)		53 (8.8)	170 (3.4)	75 (4.7)	28 (14)			

Download English Version:

<https://daneshyari.com/en/article/6208154>

Download Persian Version:

<https://daneshyari.com/article/6208154>

[Daneshyari.com](https://daneshyari.com)