



Contents lists available at ScienceDirect

International Journal of Surgery

journal homepage: www.journal-surgery.net

Original research

Evaluation of digital infra-red thermal imaging as an adjunctive screening method for breast carcinoma: A pilot study

Muffazzal Rassiwal^{a,*}, Poonam Mathur, MS^b, Rajkumar Mathur^a, Khan Farid^a, Sapna Shukla^a, P.K. Gupta^c, Beena Jain^c^a Department of Surgery, M.G.M. Medical College and M.Y. Hospital, Indore, Madhya Pradesh, India^b Department of Obstetrics & Gynaecology, M.G.M. Medical College and M.Y. Hospital, Indore, India^c LASER & Biomedical Applications, RRCAT, Indore, India

HIGHLIGHTS

- Breast thermography has been researched for over 50 years, and over 800 peer-reviewed breast thermography studies.
- Studies suggest that an abnormal thermogram is 10 times more significant than a first order family history of the disease.
- When used along with clinical examination and mammography it can significantly improve screening of breast cancer.
- The present study was motivated by the need to evaluate the technique in Indian scenario.
- Our study results confirm that thermography is a good screening modality for breast cancer.

ARTICLE INFO

Article history:

Received 14 May 2014

Received in revised form

9 September 2014

Accepted 16 October 2014

Available online 7 November 2014

Keywords:

Digital infrared thermal imaging (DITI)

Thermography

Breast cancer

Screening

Mammography

ABSTRACT

Background: Early screening plays a pivotal role in management of breast cancer. Given the socio-economic situation in India, there is a strong felt need for a screening tool which reaches the masses rather than waiting for the masses to reach tertiary centers to be screened. Digital infra-red thermal imaging (DITI) or breast thermography as a screening test offers this possibility and needs to be carefully assessed in Indian scenario. **Methods:** The study involved 1008 female patients of age 20–60 years that had not been diagnosed of cancer of breast earlier. All the subjects in this population were screened for both the breasts using DITI. Based on the measured temperature gradients (ΔT) in thermograms, the subjects were classified in one of the three groups, normal ($\Delta T \leq 2.5$), abnormal ($\Delta T > 2.5, < 3$) and potentially having breast cancer ($\Delta T \geq 3$). All those having ($\Delta T > 2.5$) underwent triple assessment that consisted of clinical examination, radiological and histopathological examination. Those with normal thermograms were subjected to only clinical examination. **Results:** Forty nine female breasts had thermograms with temperature gradients exceeding 2.5 and were subjected to triple assessment. Forty one of these which had $\Delta T \geq 3$ were proven to be having cancer of breast and were offered suitable treatment. Eight thermograms had temperature gradients exceeding 2.5 but less than 3. Most of these were lactating mothers or had fibrocystic breast diseases. As a screening modality, DITI showed sensitivity of 97.6%, specificity of 99.17%, positive predictive value 83.67% and negative predictive value 99.89%. **Conclusion:** Based on the results of this study involving 1008 subjects for screening of breast cancer, thermography turns out to be a very useful tool for screening. Because it is non-contact, pain-free, radiation free and comparatively portable it can be used in as a proactive technique for detection of breast carcinoma.

© 2014 Surgical Associates Ltd. Published by Elsevier Ltd. All rights reserved.

1. Introduction

The incidence of breast cancer is rapidly rising and it accounts for about 25% to 33% of all cancers in women in India [1]. Management of this growing problem requires effective screening

* Corresponding author.

E-mail address: muffazzalr@gmail.com (M. Rassiwal).

modalities. The widely accepted screening modality for breast cancer is the triple assessments, which comprises of self-examination or regular visits and follow ups for clinical examination, ultrasonography and mammography and biopsy [2]. However, the prevalence of these screening programs in India is rather limited because of several factors, including lack of awareness and lack of screening facilities. According to WHO's latest world health statistics, less than 5% women, aged 50–69, underwent screening by mammography in India between year 2000 and 2003 [3]. That is why over 50% of breast cancer patients in India present at late stages 3 and 4, which impedes the effective management of the cancer.

This situation might improve if a screening test is made available that is able to reach the masses rather than waiting for the masses to reach tertiary centers to be screened. Digital infra-red thermal imaging (DITI) or breast thermography as a screening test offers this possibility and needs to be carefully assessed in Indian scenario. This approach is based on detecting the heat produced due to increased blood circulation as a result of neo-angiogenesis, and metabolic changes associated with a tumor's genesis and growth. In contrast mammography, ultrasound, MRI, and other imaging tools are primarily based on differences in the scattering/transmission characteristics of tumor with respect to the normal tissue.

Infrared thermography has been in use in medical diagnostics since the 1960s, and in 1982 was approved by the US Food and Drug Administration (FDA) as an adjunctive tool for the diagnosis of breast cancer [4]. Still its use as a screening tool is rather limited at least in India. With the continued improvements in thermal imaging camera technology and the image processing capability of the present day computers, DITI looks even more attractive as a portable, non-invasive diagnostic tool. Therefore, to assess the feasibility and efficacy of DITI we conducted a pilot study at a government tertiary care centre in central India.

2. Study design

The study involved 1008 female patients of age 20–60 years that were being screened and had not been diagnosed of cancer of breast earlier. Pre examination written consent, in accordance with declaration of Helsinki, was obtained. Further, 95.15% of the subjects included were asymptomatic females, 3.67% had unilateral or bilateral mastalgia, 0.69% had breast lump but did not seek any treatment and 0.49% female complained of discharge.

3. Thermography methodology

The examination was performed in a cool carpeted room, with ambient temperature maintained around 21–25 °C, with patient

sitting disrobed up to waist in front of an infra-red camera, after at least 15 min of no physical activity. This infra-red camera (Flir, Thermo vision A-20) detects signal over the spectral range of 7.5–13 μm and had image resolution of 160×120 pixels. Its operating temperature range is from -20 °C up to 90 °C with temperature resolution ~ 0.1 °C. The software supplied with the camera was used to find areas of the breasts having different temperature gradients by generating a color-coded, processed image of the breasts showing suspicious foci. According to the magnitude of temperature gradient all images were divided in 3 classes; normal ($\Delta T \leq 2.5$), abnormal ($\Delta T > 2.5, < 3$) and potentially having breast cancer ($\Delta T \geq 3$). All those having $\Delta T > 2.5$ underwent triple assessment that consisted of clinical examination, radiological and histopathological examination. Those with normal thermograms were subjected to only clinical examination.

4. Results

Out of the 1008 females screened, $\Delta T > 2.5$ was observed for 49 cases. Out of these 41 subjects had ($\Delta T \geq 3$). On subsequent examination by triple assessment all these breasts (having $\Delta T \geq 3$) were found to be the cases of breast cancer. Eight cases had $\Delta T > 2.5$ but less than 3. Of these 3 were lactating mothers, 5 had fibrocystic disease (fibroadenosis-4, fibroadenoma-1) as shown in Table 1. Those with insignificant thermographic abnormality were excluded for malignancy [4]. Some representative thermograms with $\Delta T \geq 3$ indicating of a suspicious foci suggesting unilateral breast cancer are shown in Fig. 1a,b & c (details are given in figure legends). Thermogram, in Fig. 2 with $\Delta T > 2.5$ and < 3 shows lactating engorged breast with bilateral symmetrical vascular signs. Thermograms in Fig. 3, with $\Delta T \leq 2.5$, shows bilateral fibroadenoma with normal thermal image suggestive of a benign pathology and ruling out malignancy which otherwise would have needed an invasive procedure like fine needle aspiration cytology or biopsy.

Based on the subjects included in this study, sensitivity of breast thermography to detect malignancy came out to be 97.6% with 1 case being false negative. Specificity was 99.17%, positive predictive value came out to be 83.67% and negative predictive value 99.89% as shown in Table 2.

5. Discussion

The risk of getting breast cancer has tripled over the last half-century due to change in life style and other factors [5]. Therefore it is necessary to detect breast cancer early for better management of this growing menace. For decades, breast self-examination and mammography have been used as the screening techniques of choice. However, there have been concerns about the safety of mammography [6–8]. These concerns have led to a recent revision

Table 1
Details about the subject examined in study and their outcomes.

No. of subjects	Thermography	Triple assessments	Remarks
959	$\Delta T < 2.5$	<ul style="list-style-type: none"> Clinical examination normal 	Negative for malignancy
8	$\Delta T > 2.5$ but < 2.9	<ul style="list-style-type: none"> 5 patients had suspicious lump on clinical examination. All 8 mammograms were normal. Biopsy confirmed fibrocystic disease in 5 patients.(fibroadenosis-4, fibroadenoma-1) 	Negative for malignancy
41	$\Delta T > 3$	<ul style="list-style-type: none"> All patients had clinically palpable lump on examination. 38 patients had abnormal mammography Biopsy: 25 patients had Infiltrating Ductal Carcinoma 14 patients had Ductal Carcinoma in situ 1 patient had phylloides tumour 1 patient had inflammatory carcinoma of breast 	3 patients had normal mammography but were found to be infiltrating ductal carcinoma.

Download English Version:

<https://daneshyari.com/en/article/6251865>

Download Persian Version:

<https://daneshyari.com/article/6251865>

[Daneshyari.com](https://daneshyari.com)