

## Author's Accepted Manuscript

Multi-axial damage and failure of medical grade carbon fibre reinforced PEEK laminates: experimental testing and computational modelling

Elizabeth Anne Gallagher, Steven Lamorinière,  
Patrick McGarry



PII: S1751-6161(18)30310-2  
DOI: <https://doi.org/10.1016/j.jmbbm.2018.03.015>  
Reference: JMBBM2724

To appear in: *Journal of the Mechanical Behavior of Biomedical Materials*

Received date: 8 December 2017  
Revised date: 7 March 2018  
Accepted date: 12 March 2018

Cite this article as: Elizabeth Anne Gallagher, Steven Lamorinière and Patrick McGarry, Multi-axial damage and failure of medical grade carbon fibre reinforced PEEK laminates: experimental testing and computational modelling, *Journal of the Mechanical Behavior of Biomedical Materials*, <https://doi.org/10.1016/j.jmbbm.2018.03.015>

This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting galley proof before it is published in its final citable form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

Multi-axial damage and failure of medical grade carbon fibre reinforced PEEK laminates:  
experimental testing and computational modelling

Elizabeth Anne Gallagher<sup>a</sup>, Steven Lamorinière<sup>b</sup>, Patrick McGarry<sup>a</sup>

**a:** Discipline of Biomedical Engineering, National University of Ireland Galway, University Road, Galway, Ireland

**b:** Invibio Ltd., Hillhouse International, Thornton-Cleveleys, FY5 4QD, United Kingdom

Corresponding Author: Dr. Patrick McGarry  
College of Engineering and Informatics  
National University of Ireland Galway  
Galway, Ireland  
Email: patrick.mcgarry@nuigalway.ie

**Abstract**

Orthopaedic devices using unidirectional carbon fibre reinforced poly-ether-ether-ketone (PEEK) laminates potentially offer several benefits over metallic implants including: anisotropic material properties; radiolucency and strength to weight ratio. However, despite FDA clearance of PEEK-OPTIMA™ Ultra-Reinforced, no investigation of the mechanical properties or failure mechanisms of a medical grade unidirectional laminate material has been published to date, thus hindering the development of first-generation laminated orthopaedic devices. This study presents the first investigation of the mechanical behaviour and failure mechanisms of PEEK-OPTIMA™ Ultra-Reinforced. The following multi-axial suite of experimental tests are presented: 0° and 90° tension and compression, in-plane shear, mode I and mode II fracture toughness, compression of ±45° laminates and flexure of 0°, 90° and ±45° laminates. Three damage mechanisms are uncovered: (1) inter-laminar delamination, (2) intra-laminar cracking and (3) anisotropic plasticity. A computational damage and failure model that incorporates all three damage mechanisms is developed. The model accurately predicts the complex multi-mode failure mechanisms observed experimentally. The ability of a model to predict diverse damage mechanisms under multiple loading directions conditions is critical for the safe design of fibre reinforced laminated orthopaedic devices subjected to complex physiological loading conditions.

Download English Version:

<https://daneshyari.com/en/article/7207053>

Download Persian Version:

<https://daneshyari.com/article/7207053>

[Daneshyari.com](https://daneshyari.com)