

Author's Accepted Manuscript

A Review of Hydroxyapatite-based Coating Techniques: Sol-gel and Electrochemical Depositions on Biocompatible Metals

R.I.M. Asri, W.S.W. Harun, M.A. Hassan, S.A.C. Ghani, Z. Buyong



PII: S1751-6161(15)00449-X
DOI: <http://dx.doi.org/10.1016/j.jmbbm.2015.11.031>
Reference: JMBBM1704

To appear in: *Journal of the Mechanical Behavior of Biomedical Materials*

Received date: 17 September 2015
Revised date: 26 November 2015
Accepted date: 30 November 2015

Cite this article as: R.I.M. Asri, W.S.W. Harun, M.A. Hassan, S.A.C. Ghani and Z. Buyong, A Review of Hydroxyapatite-based Coating Techniques: Sol-gel and Electrochemical Depositions on Biocompatible Metals, *Journal of the Mechanical Behavior of Biomedical Materials* <http://dx.doi.org/10.1016/j.jmbbm.2015.11.031>

This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting galley proof before it is published in its final citable form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

A Review of Hydroxyapatite-based Coating Techniques: Sol-gel and Electrochemical Depositions on Biocompatible Metals

R.I.M. Asri¹, W.S.W. Harun², M.A. Hassan², S.A.C. Ghani², Z. Buyong³

¹Institute of Postgraduate Studies, Universiti Malaysia Pahang, Lebuhraya Tun Razak,
26300 Gambang, Kuantan, Pahang, Malaysia

²Human Engineering Group, Faculty of Mechanical Engineering,
Universiti Malaysia Pahang, 26600 Pekan, Pahang, Malaysia

³Kuliyah of Medicine, International Islamic University Malaysia,
25200 Kuantan, Pahang, Malaysia

Corresponding author: sharuzi@ump.edu.my

New promising techniques for depositing biocompatible hydroxyapatite -based coatings on biocompatible metal substrates for biomedical applications have continuously been exploited for more than two decades. Currently, various experimental deposition processes have been employed. In this review, the two most frequently used deposition processes will be discussed: A sol-gel dip coating and an electrochemical deposition. This study deliberates the surface morphologies and chemical composition, mechanical performance and biological responses of sol-gel dip coating as well as the electrochemical deposition for two different sample conditions, with and without coating. The review shows that sol-gel dip coatings and electrochemical deposition were able to obtain the uniform and homogeneous coating thickness and high adherent biocompatible coatings even in complex shapes. It has been accepted that both coating techniques improve bone strength and initial osseointegration rate. The main advantages and limitations of those techniques of hydroxyapatite-based coatings are presented. Furthermore, the most significant challenges and critical issues are also highlighted.

Download English Version:

<https://daneshyari.com/en/article/7208178>

Download Persian Version:

<https://daneshyari.com/article/7208178>

[Daneshyari.com](https://daneshyari.com)