

## Accepted Manuscript

Title: Hyphenation of Capillary Electrophoresis to Inductively Coupled Plasma Mass Spectrometry with a Modified Coaxial Sheath-flow Interface

Authors: Hannah U. Holtkamp, Stuart J. Morrow, Mario Kubanik, Christian G. Hartinger



PII: S0021-9673(18)30619-8  
DOI: <https://doi.org/10.1016/j.chroma.2018.05.036>  
Reference: CHROMA 359404

To appear in: *Journal of Chromatography A*

Received date: 20-2-2018  
Revised date: 14-5-2018  
Accepted date: 15-5-2018

Please cite this article as: Hannah U.Holtkamp, Stuart J.Morrow, Mario Kubanik, Christian G.Hartinger, Hyphenation of Capillary Electrophoresis to Inductively Coupled Plasma Mass Spectrometry with a Modified Coaxial Sheath-flow Interface, *Journal of Chromatography A* <https://doi.org/10.1016/j.chroma.2018.05.036>

This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

# Hyphenation of Capillary Electrophoresis to Inductively Coupled Plasma Mass Spectrometry with a Modified Coaxial Sheath-flow Interface

Hannah U. Holtkamp,<sup>a</sup> Stuart J. Morrow,<sup>a</sup> Mario Kubanik,<sup>a</sup> and Christian G. Hartinger<sup>a,\*</sup>

<sup>a</sup> University of Auckland, School of Chemical Sciences, Private Bag 92019, Auckland 1142, New Zealand. <http://hartinger.auckland.ac.nz/>.

\* Corresponding author. E-mail address: [c.hartinger@auckland.ac.nz](mailto:c.hartinger@auckland.ac.nz) (C. G. Hartinger); Tel +64 9 3737 599 ext 83220.

Electronic Supplementary Information (ESI) available: ICP-MS settings, additional data collected during the method optimization, LOD comparisons.

## Highlights

- Adaptation of a coaxial sheath-flow interface for CE–ICP-MS hyphenation
- Comparison of the analytical performance to a commercially available interface
- Optimization of key parameters for CE–ICP-MS hyphenation
- proof-of-principle experiments demonstrated very good analytical performance

Download English Version:

<https://daneshyari.com/en/article/7607794>

Download Persian Version:

<https://daneshyari.com/article/7607794>

[Daneshyari.com](https://daneshyari.com)