



# Comparison of the behavior of CHO cells during cultivation in 24-square deep well microplates and conventional shake flask systems



Kirti Chaturvedi\*, Susan Y. Sun, Thomas O'Brien, Yan J. Liu, James W. Brooks

BD Biosciences – Advanced Bioprocessing, Cockeysville, MD, USA

## ARTICLE INFO

### Article history:

Available online 6 May 2014

### Keywords:

CHO cells  
Duetz sandwich-cover system  
Batch  
Fed batch  
Cell culture  
Deep well plates

## ABSTRACT

In biopharmaceutical production, the optimization of cell culture media and supplementation is a vital element of process development. Optimization is usually achieved through the screening of multiple media, feed and feeding strategies. However, most screening is performed in shake flasks, which makes the screening process very time consuming and inefficient. The use of small scale culture systems for the screening process can aid in the ability to screen multiple formulations during process development. In order to assess the suitability of 24 deep well (24DW) plates with the Duetz sandwich-covers as a small scale culture system for process development, we have tested growth and production performance of CHO cells in 24DW plates and conventional shake flask cultures. Multiple studies were performed to assess well-to-well and plate-to-plate variability in 24DW plates. Additional studies were performed to determine the applicability of 24DW plates for cell culture medium and supplement screening in batch and fed batch processes. Cultures in 24DW plates exhibited similar kinetics in growth, viability and protein production to those cultured in shake flasks, suggesting that 24DW plates with Duetz sandwich-covers can be effectively used for high throughput cell culture screening.

© 2014 The Authors. Published by Elsevier B.V. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/3.0/>).

## 1. Introduction

The demand for antibodies and other glycoproteins has increased rapidly due to their importance as therapeutic agents. Production of these biologics is in large part reliant on mammalian cell culture systems. To meet the demand, various strategies are employed to increase production by enhancing cell performance, specifically improving bioproduction titers. Some of the most effective approaches to improve the cell culture process include media optimization and batch and fed batch supplementation. Development of these strategies involves intensive screening of components, mixtures and various feed schemes. Due to the difficulty of examining a large number of components at different concentrations or screening a large number of potential supplements, these types of studies cannot be efficiently performed using

conventional bioreactor, shake flask, or spin tube cultures. Use of small scale cell culture systems, like multi-well plates can enable the exploration of a wider range of bioprocess operating conditions in a more efficient manner.

In general, multi-well plate cell cultures are performed in static or shaking formats in shallow well (SW) or deep well (DW) plates in a small culture volume for elongated time periods. Shaking plates are more suitable for suspension cell culture; however, the loss of media due to evaporation in outer wells (edge effects) poses a significant problem [1]. Edge effects result in well-to-well and plate-to-plate variability in cultures. Some laboratories have overcome the edge effects and evaporation issues by not using the outer wells of multi-well plates for cell-based assays. Omission of outer wells decreases analysis throughput by 38% and 66% in 96 well and 24 well plates, respectively. Other approaches to prevent evaporation include the use of self-adhesive plate seals, which are designed to maintain uniform gas exchange while keeping cultures sterile. A study of various commercially available adhesive plate seals showed multiple disadvantages and categorized them into two groups: (1) plate seals in which volume preservation is relatively low, but oxygen transfer is comparable to that of unsealed plates, and (2) plate seals in which volume preservation is high, but oxygen exchange is slower [2]. Therefore, adhesive plate seals may

*Abbreviations:* CD, chemically defined; 24DW, 24 deep well; VCD, viable cell density; CHO cells, Chinese Hamster Ovarian cells; %CV, percent coefficient of variation.

\* Corresponding author at: BD Advanced Processing, 250 Schilling Circle, Cockeysville, MD 21030, USA. Tel.: +1 410 773 6415.

E-mail address: [kirti.chaturvedi@bd.com](mailto:kirti.chaturvedi@bd.com) (K. Chaturvedi).

<http://dx.doi.org/10.1016/j.btre.2014.04.001>

2215-017X/© 2014 The Authors. Published by Elsevier B.V. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/3.0/>).

not adequately fulfill the requirements for cell culture screening in SW and DW multi-well plates. In addition to evaporation and oxygen transfer issues, culture volumes in most of the 24SW, 96SW, and 96DW plates ranges from 0.7 mL to 1 mL. These lower volumes are usually insufficient for multiple sampling and fed batch processes. Moreover, for suspension cultures, plates are being kept on shaking platform to keep cells in suspension. It is often challenging to keep plates secure on a regular platform while shaking.

These problems can be adequately addressed using a closure system for multi well plates as described by Duetz et al. [3]. The Duetz sandwich-cover system is a shaking multi-well plate based system that consists of specialized multi-well plate sandwich covers and clamps to hold the closures in place. The sandwich cover is an autoclavable stainless steel lid containing layers of filters and silicon sealing that provides a positive seal on each well of the plate to promote efficient gas transfer. The headspace refreshment rate of each individual well is controlled by a small hole in the silicone layer above the center of each well and contamination is prevented using autoclavable gas permeable filters [3,4]. The system provides a headspace refreshment rate of 0.1–2 working volumes per minute in orbital shakers, permitting oxygen concentrations of at least 18% (v/v), even when oxygen uptake rates are as high as 40 mmol O<sub>2</sub>/L/h. Evaporation at these conditions is kept at a minimum (as low as 2% per day), which permits longer culture times [5]. The clamps assure that the plates and sandwich covers are clamped together tightly and the individual wells are hermetically sealed. The cover clamp can be mounted onto a variety of regular orbital shaking platforms. The Duetz sandwich-cover system can be used with various multi well plates, including 24DW plates. A 24DW plate can hold up to four milliliter culture volume as compare to one milliliter in any other multi-well plates, which enables multiple samplings on various days of culture. This Duetz sandwich-cover system has been used for bacterial cell culture to maintain oxygen transfer and reduce evaporation [6]. The system has also been used for batch and fed batch culture studies with hybridoma and Chinese Hamster Ovary (CHO) cell lines in polystyrene 24 round well plates [7,8].

In this study, we have evaluated the Duetz sandwich-cover system for CHO cell screening studies. CHO cells are the most commonly used mammalian cells for production of biopharmaceuticals. We have tested monoclonal antibody (mAb) producing CHO cell lines in 24DW plates and compared performance to conventional shake flask cultures. Initially, a series of experiments were performed to assess well-to-well and plate-to-plate variability in the 24DW plate. Additional studies were performed to determine the application of the Duetz sandwich-cover system for cell culture medium and supplement screening in batch and fed batch processes. Multiple CHO cell lines were used to ensure that scalability to shake flask culture was not cell line specific. Overall, 24DW plates gave similar kinetics in growth, viability and protein production to those cultured in shake flasks, demonstrating a potential application of 24DW plates with the Duetz sandwich-cover system in high throughput screening for cell culture process development.

## 2. Materials and methods

### 2.1. Cell lines, cell culture media and supplements

Studies were carried out using five proprietary mAb producing CHO cell lines. Cell lines 1, 4 and 5 were derived from the CHO K1 host, while cell lines 2 and 3 were derived from CHO DHFR and GS CHO hosts, respectively. Each cell line, except CHO line 4, was cultured in its optimal basal chemically defined (CD) medium for maintenance and batch and fed batch studies. Cell line 4 was maintained in a peptone containing medium, while batch and fed batch studies were performed in CD medium. Cell culture

media utilized were: BD Select CHO and BD Select CD1000 (BD Advanced Bioprocessing), CDM4CHO (Thermo Fisher Scientific), and EX CELL CD CHO3 (SAFC). Feeds and media supplements utilized were: TC Yeastolate (TCY) and Proteose Peptone 3 (PP3) (BD Advanced Bioprocessing) and CD Cell Boost 6 (Thermo Fisher). The Duetz sandwich-cover system and 24DW plates were obtained from EnzyScreen BV (Haarlem, Netherlands). The system includes 24DW plates (40 mm deep, pyramid bottom, volume 11 mL/well, transparent polystyrene plates), sandwich covers (CR1224a) and clamps (CR1700).

### 2.2. Bioassays

Bioassay cultures were seeded at identical seeding density in 24DW plates and shake flasks. Culture volume was 3 mL and 50 mL in 24DW plates and shake flasks, respectively. Plate and shake flask cultures were incubated on a shaking platform in 5% CO<sub>2</sub> at 37 °C. The shaking speed for plates was 300 rpm, while the shaking speed for flasks was 125 rpm. The orbital diameter of the shaking platform was 25 mm for both plates and shake flasks. For 24DW plates, 300 µl samples were collected from the same well on various days of culture. These samples were diluted 2–3 times with PBS (Cellgro®) for assessment of cell growth (Viable cell density; VCD), viability and protein production. VCD and viability were determined using a Vi-CELL® (Beckman Coulter) and protein production was measured using a ForteBio Octet® (Pall Life Sciences). For batch culture studies, peptones were added to basal media on Day 0. Fed batch cultures were fed with CD feeds on alternate days starting on Day 0 of the culture. Peptone titration studies were performed to test the effect of various concentrations of peptones on growth and production of CHO cells in a batch culture.

### 2.3. Statistical analyses

Minitab 16 software (Minitab Inc) was used to generate multi-variate charts and to perform other statistical analyses. Correlation analysis was performed to determine growth and production performance relationship between shake flask and 24DW plates. The Pearson correlation coefficient is a measure of linear association between two variables. Values of the correlation coefficient are always between –1 and +1. A correlation coefficient of +1 indicates that two variables are perfectly related in a positive linear sense. Two-Way ANOVA was used to assess effect of plates and peptone titration in plate-to-plate variation study.

## 3. Results

### 3.1. Comparison of 24DW plate with Duetz sandwich-cover system and conventional shake flask system

Four CHO lines were grown concurrently in their respective optimal base media in 24DW plates with Duetz sandwich-covers and in conventional shake flasks. Samples were collected on various days of culture and operational parameters for each culture system were as described in Material and Method. Regardless of medium and cell line, growth kinetics and mAb production of CHO cell lines in 24DW plates was comparable to those grown in shake flasks (Fig. 1). As shown in Fig. 2, cell viabilities were maintained above 80% on Day 7 of culture for all cell lines in both shake flask and 24DW plates. These results indicate that 24DW plates may simulate the performance and dynamics of shake flask and can be used for cell culture process development studies.

Download English Version:

<https://daneshyari.com/en/article/870708>

Download Persian Version:

<https://daneshyari.com/article/870708>

[Daneshyari.com](https://daneshyari.com)