

Letter to the Editor



Comparative Analysis of the Multi-Frequency Bio-impedance and Dual-energy X-ray Absorptiometry on Body Composition in Obese Subjects

WANG Zheng He, YANG Zhong Ping, WANG Xi Jie, DONG Yan Hui, and MA Jun[#]

To examine accuracy of body composition predicted by the Multi-Frequency Bioelectric Impedance Analysis (MF-BIA) compared with the Dual-energy X-ray Absorptiometry (DXA) in adults with obese. We measured body composition of 749 adults with obese both by the MF-BIA and DXA. The Lin's concordance correlation and the Bland-Altman plots were used to examine the consistency. The concordance correlation coefficient of %BF between the MF-BIA and DXA in men and women was 0.560, and 0.669, respectively. Compared with the DXA, the MF-BIA significantly underestimated %BF by 4.33% in men ($P < 0.001$), however overestimated %BF by 0.50% in women ($P < 0.001$). After corrected by the correction equations established in this study, the differences were significantly decreased. Therefore, the MF-BIA (TANITA MC-180) may need to be corrected in estimating body composition for adults with obese.

Body composition information, including the percentage of body fat (%BF) and fat free mass (FFM), is widely used to evaluate the nutrition condition and the risk factors related to obesity. In the last decades, with the rapidly rising prevalence of obesity worldwide, there is a more and more urgent need for a simple and rapid method to assess accurately the body composition for population with obese.

Several technologies have been developed to measure the body composition information, such as the underwater weighting method^[1] and the dual-energy X-ray absorptiometry (DXA). The DXA was a new developed reference method to measure the body composition^[2]. However, due to the inaccessibility and the high cost of the equipment, DXA was hardly used in the large epidemiological survey. Thus, general methods such as the bioelectrical impedance analysis (BIA) were still the most common methods in epidemiological

studies^[3-4]. Since the BIA was firstly validated a commercially available by Lukaski, et al.^[5], it has been widely used in the clinical practices, epidemiological surveys, and sports medicine. However, studies found that the accuracy were good for non-obese individuals, but not for obese individuals^[6-7].

Due to the BIA manufacturers often based on the population with certain race and nutrition condition to establish the prediction equations, which might not apply to all individuals, such as individuals with obese. However, the BIA was more commonly applied in individuals with overweight or obese than that with normal weight to measure body fat percentage. Several studies have found that the MF-BIA significantly underestimated %BF of obese adults^[6], but the sample sizes of these studies were very limited. Thus, whether the MF-BIA could accurately predict the body composition in Chinese adults with obese was still unclear. Therefore, we performed the parallel measurements of body composition (including %BF and FFM) by the MF-BIA and DXA in adults with obese from Beijing, China, to evaluate the accuracy between the two methods and provide correction equations to more accurately measure the body composition in Chinese adults with obese by the MF-BIA.

Study Population A total of 749 participants were voluntarily recruited through recruitment poster from Beijing, China. According to the inclusion and exclusion criteria for participants in the study design, the participants aged 25-55 years who lived in Beijing for more than 1 year, with BMI ≥ 28.0 kg/m², and without suffering from severe organic diseases were included in the current study. The medical ethical committee of Peking University approved the study (IRB00001052-13086) and all procedures were carried out with adequate understanding and written consent of all the participants.

doi: 10.3967/bes2018.008

Institute of Child and Adolescent Health, School of Public Health, Peking University Health Science Center, Beijing 100191, China

Measurements of the MF-BIA and DXA All the measurements were performed after ≥ 8 -h fasting. The TANITA MC-180 Body Composition Analyzer (TANITA, MC-180, Japan) was used to measure body composition, including the %BF and FFM. Participants were asked to stand on the Body Composition Analyzer with bare feet, remove all the accessories, and make sure there's no metal in the body. Body composition was measured by the prediction equations of manufacturer within the Analyzer. After assessed by the MF-BIA, the DXA method was also used to assess body composition within 2 hours. DXA measurements were conducted in the Institute of Sports Medicine, Peking University Third Hospital by professional technician using the General Electric Lunar Prodigy Scanner (GE Healthcare, Lunar iDXA ME+ 210205, USA) with standard model (The dose of X-ray was 0.4 LGy, and the supporting current was 0.15 A) in a supine position. All the metal objects must be removed from participants to ensure accuracy.

Statistical Analysis All the statistical analyses were performed by the SPSS 17.0 for Windows statistical package (SPSS, Chicago, IL). Independent sample *t*-test was used to compare the gender differences in age, height, weight, and body composition including %FM and FFM. Paired sample *t*-test was used to examine differences of body composition parameters between the MF-BIA and the DXA. The Lin's concordance correlation and Bland-Altman plot were used to test the accuracy between the two methods. To establish correction equations of the MF-BIA (TANITA, MC-180), we randomly divided all participants into two groups, the first group ($n = 374$) was used to establish correction equations using the simple linear

regression model with gold standard of DXA, stratified by sex. The second group ($n = 375$) was used to evaluate the effect of the correction equations through the Lin's concordance correlation and paired sample *t*-test. A $P < 0.05$ for two-sided was considered statistically significant.

Consistency Analysis between the MF-BIA and DXA

The results of consistency analysis between the MF-BIA and DXA for Chinese adults with obese were showed in Table 1. The average %BF measured by the MF-BIA was significantly lower (4.33%) than that measured by the DXA method among men ($P < 0.001$). However, it was significantly higher (0.50%) than the DXA among women ($P < 0.001$). The CCC between %BF measured by the MF-BIA and DXA for men and women were 0.560 and 0.669, respectively. The average of FFM (kg), measured by the MF-BIA, is significantly higher 3.8% than the DXA among men ($P < 0.001$), however it is significantly lower 5.1% than the DXA among women ($P < 0.001$). The CCC between the FFM measured by the MF-BIA and DXA among men and women were 0.866 and 0.632, respectively.

Bland-Altman Plot Analyses As shown in Figure 1. The mean %BF for men is 36.09% (Table 1), the 2SD (standard error) LOA (limit of agreement) is around 12% (Figure 1A), thus the LOA is around 38%. The mean %BF for women is 44.00% (Table 2), the 2SD LOA is around 13% (Figure 1B), and thus the LOA is around 30%. The mean FFM for men is 60.16 kg (Table 2), the 2SD LOA (Limit of Agreement) is about 14 kg (Figure 1C), and thus the LOA are around 23%. The mean FFM for women is 44.36 kg (Table 2), the 2SD LOA is about 10 kg (Figure 1D), and thus the LOA is around 23%. These results clearly showed that the agreement between the MF-BIA and DXA was poor in individuals with obese.

Table 1. The Results of %BF and Fat Free Mass in kg Predicted by the MF-BIA and DXA method before and after correction

	Men ($n = 178$)				Women ($n = 196$)			
	MF-BIA	DXA	<i>d</i>	ρ_c	MF-BIA	DXA	<i>d</i>	ρ_c
Before correction								
%BF	31.75 \pm 4.83	36.09 \pm 4.41	-4.33 \pm 2.88	0.560	44.50 \pm 4.44	44.00 \pm 3.38	0.50 \pm 3.19	0.669
FFM	62.43 \pm 6.08	60.16 \pm 7.11	2.27 \pm 3.57	0.866	42.11 \pm 3.04	44.36 \pm 4.83	-2.25 \pm 2.96	0.632
After correction								
%BF	36.18 \pm 3.25	36.10 \pm 4.50	0.09 \pm 2.62	0.780	44.13 \pm 2.13	44.00 \pm 3.30	0.09 \pm 2.41	0.627
FFM	60.50 \pm 6.02	60.50 \pm 6.64	0.01 \pm 2.67	0.911	43.05 \pm 3.69	44.50 \pm 4.73	-1.46 \pm 2.58	0.770

Note. %BF, percentage of body fat; FFM, fat free mass; MF-BIA, multi-frequency bio-impedance; DXA, dual energy X-ray absorptiometry; *d*, mean difference of MF-BIA from DXA; ρ_c , concordance correlation coefficient.

Download English Version:

<https://daneshyari.com/en/article/8817480>

Download Persian Version:

<https://daneshyari.com/article/8817480>

[Daneshyari.com](https://daneshyari.com)