

#### Available online at www.sciencedirect.com



JOURNAL OF CHROMATOGRAPHY A

Journal of Chromatography A, 1071 (2005) 55–58

www.elsevier.com/locate/chroma

#### Short communication

# Increasing the accuracy of determination of $n_{\rm C}/n_{\rm H}$ ratios by gas chromatography—atomic emission detection

Elena S. Chernetsova<sup>a,\*</sup>, Alexander I. Revelsky<sup>a</sup>, Dupont Durst<sup>b</sup>, Tim G. Sobolevsky<sup>a</sup>, Igor A. Revelsky<sup>a</sup>

<sup>a</sup> Analytical Chemistry Division, Chemistry Department, Moscow State University, Leninskie Gory 1 Building 3, 119992 Moscow, Russian Federation
<sup>b</sup> U.S. Army Edgewood Laboratories, Aberdeen Proving Ground, MD, USA

#### **Abstract**

The influence of oxygen content in helium on the accuracy of  $n_{\rm C}/n_{\rm H}$  ratio determination for model mixtures of aliphatic and polyaromatic hydrocarbons and polychlorinated biphenyls was studied. The best accuracy was achieved at the oxygen content ca. 9%, which was the maximal possible oxygen content in helium for this GC–atomic emission detection (helium flow rate was 25 ml min<sup>-1</sup>). Using the maximal oxygen flow in plasma the  $n_{\rm C}/n_{\rm H}$  ratio determination accuracy improvement was accompanied by 10-fold increase in detection limit. © 2004 Elsevier B.V. All rights reserved.

Keywords: Gas chromatography-atomic emission detection;  $n_{\rm C}/n_{\rm H}$  ratios; Aliphatic hydrocarbons

#### 1. Introduction

The developing of empirical formulae determination method is one of the possible ways to improve the reliability of identification of unknowns. From this viewpoint, attractive is gas chromatography with atomic emission detection, because, providing full atomization of analytes, the atomic emission detection (AED) signal should be independent of their structures. Therefore, empirical formulae could be determined using only one mixture component as a reference for other components.

In the large majority of publications it has been shown that the AED response is dependent on analyte structure and concentration and plasma conditions [1–4]. The requirement of accurate determination of elemental ratios is that the reference compound has a structure, retention time, concentration and elemental composition similar to compound being determined [2,5–7]. However, instrumental parameters influencing the accuracy of elemental ratio determination by GC–AED have not been thoroughly investigated, and usually

the instrument is operated at gas pressures recommended by manufacturer without indication of the respective flow rates.

The GC-AED system developed by Hewlett-Packard is most widely used nowadays. In order to improve the chromatographic peaks shape, increase the plasma electron density and prevent carbon deposition on discharge tube walls, such reagent gases as oxygen, hydrogen, and/or nitrogen-methane (9:1) mixture are added to the helium plasma, depending on element wavelengths to be registered [8,9]. Oxygen is used as a reagent gas for simultaneous C and H element chromatograms registration.

Thus, we have decided to study the influence of oxygen content in helium on the accuracy of  $n_{\rm C}/n_{\rm H}$  ratio determination for model mixtures of aliphatic and polyaromatic hydrocarbons and polychlorinated biphenyls.  $n_{\rm C}/n_{\rm H}$  ratios were determined for all model compounds using tridecane as a reference compound.

#### 2. Experimental

#### 2.1. Chemicals

Undecane, dodecane, tridecane, tetradecane, pentadecane, and hexadecane were purchased from Aldrich (Milwaukee,

<sup>\*</sup> Corresponding author. Tel.: +7 095 9394318; fax: +7 095 9394675. *E-mail address:* chernetsova@environment.chem.msu.ru (E.S. Chernetsova).

WI, USA). Naphthalene ( $C_{10}H_8$ ), acenaphthene ( $C_{12}H_{10}$ ), fluorene ( $C_{13}H_{10}$ ), pyrene ( $C_{16}H_{10}$ ), and chrysene ( $C_{18}H_{12}$ ) were purchased as  $2\times 10^{-4}\,\mathrm{g/\mu l}$  solutions in acetonitrile from Ecros (St. Petersburg, Russia). 2-Chlorobiphenyl, 2,3-dichlorobiphenyl, 2,4,5-trichlorobiphenyl, and 2,2',4,4'-tetrachlorobiphenyl were purchased as  $1\times 10^{-7}\,\mathrm{g/\mu l}$  mixture in acetone from Ultra Scientific (North Kingstown, RI, USA). Hexane, acetonitrile and acetone were supplied from Lecbiopharm (Moscow, Russia).

Helium (99.9999%), nitrogen (99.999%), oxygen (99.999%), hydrogen (99.9999%), and nitrogen/methane mixture (9:1) (99.99%) were obtained from PromGasService (Moscow, Russia).

The stock solutions containing  $5 \times 10^{-5}$  g/µl per individual hydrocarbon were prepared by dissolving 50 mg of each hydrocarbon in 1 ml of *n*-hexane. The model mixtures were prepared by dilution and/or mixing of the stock solutions.

The model mixtures of polyaromatic hydrocarbons and polychlorinated biphenyls were prepared in acetone and acetonitrile, respectively. Tridecane was added as an internal standard.

#### 2.2. GC-AED

A microwave-induced plasma atomic emission detector, model "G2350A" (Agilent, Waldbronn, Germany), was coupled to a "6890N" gas chromatograph. The GC was equipped with a "7683" automatic sampler and a split-splitless capillary injection port. For instrument control and operation G2370AA Chemstation with the G2360AA GC-AED software was used. Chromatographic separation was performed on a  $25 \,\mathrm{m} \times 0.32 \,\mathrm{mm}$  i.d. HP-1 100% polydimethylsiloxane capillary column (Agilent) with 0.17 µm film thickness. Sample volumes of 1 µl were injected by a 10 µl syringe of the automatic sampler in splitless mode (injection port temperature 280 °C). The flow rate of carrier gas (helium) was 1.5 ml min<sup>-1</sup>. Model compounds were fully separated applying temperature programming. Following injection, the solvent was vented for 4 min. The transfer line and detector cavity were operated at 280 °C. Carbon and hydrogen were registered at 496 and 486 nm, respectively. Oxygen was used as a reagent gas. The helium and oxygen flow rates through the plasma were calculated using the flow rate values measured at the vent lines of the AED system by a bubble meter at different oxygen pressures. The range of oxygen flow rates we investigated was 0-2.5 ml min<sup>-1</sup> (corresponded to pressures 0–6.89 bar). The oxygen pressure at the recommended by the manufacturer instrument operation mode was 1.01 bar, the corresponding oxygen content in plasma was ca. 1.5%.

### 3. Results and discussion

Providing the independence of detector C and H responses on structure of the analytes,  $n_{\rm C}/n_{\rm H}$  ratio can be calculated for unknown compound by use of the equation:

$$\frac{n_{\rm C,unknown}}{n_{\rm H,unknown}} = \frac{S_{\rm C,unknown}}{S_{\rm H,unknown}} \frac{n_{\rm C,reference}}{n_{\rm H,reference}} \frac{S_{\rm H,reference}}{S_{\rm C,reference}},$$

where  $n_{\text{C,unknown}}/n_{\text{H,unknown}}$  and  $n_{\text{C,reference}}/n_{\text{H,reference}}$  are the ratios of the numbers of C and H atoms in the molecule of unknown and reference compound, respectively;  $S_{\text{C,unknown}}/S_{\text{H,unknown}}$  and  $S_{\text{H,reference}}/S_{\text{C,reference}}$  are the ratios of peak areas registered on C and H channels for unknown and reference compound, respectively.

The relative errors of calculated  $n_{\rm C}/n_{\rm H}$  ratios can be evaluated as:

$$\Delta (\%) = \frac{|(n_{\text{C,calc.}}/n_{\text{H,calc.}}) - (n_{\text{C,real}}/n_{\text{H,real}})|}{n_{\text{C,real}}/n_{\text{H,real}}} \times 100\%,$$

where  $n_{\rm C,real}/n_{\rm H,real}$  and  $n_{\rm C,calc.}/n_{\rm H,calc.}$  are real and calculated  $n_{\rm C}/n_{\rm H}$  ratios, respectively.

## 3.1. $n_C/n_H$ ratio determination for model aliphatic hydrocarbons

We have studied the accuracy of  $n_C/n_H$  ratio determination for the mixtures containing  $5 \times 10^{-7}$ ,  $1 \times 10^{-7}$ ,  $5 \times 10^{-8}$ ,  $1 \times 10^{-8}$ ,  $5 \times 10^{-7}$ , and  $5 \times 10^{-9}$  g/ $\mu$ l per component of aliphatic hydrocarbons  $C_{11}$ - $C_{16}$ , using tridecane as a reference for other components, and found that the AED response was independent on compound structure, when oxygen content in helium was equal to 1.5% (which was set according to manufacturer recommendations; hereinafter we will name it as "standard oxygen content"). As an example, the results of  $n_C/n_H$  ratio determination at the standard oxygen content in helium are given in Table 1 (the concentration of  $C_{11}$ - $C_{16}$  hydrocarbons was  $5 \times 10^{-7}$  g/ $\mu$ l).

As it is seen from Table 1, in the standard instrument operation mode no significant  $n_{\rm C}/n_{\rm H}$  ratio determination accuracy dependence on structure of  $C_{11}$ – $C_{16}$  hydrocarbons was observed.

Aliphatic hydrocarbons  $n_{\rm C}/n_{\rm H}$  ratio determination results at the standard oxygen content in helium (1.5%) (n=5, P=0.95; components concentration  $5 \times 10^{-7}$  g/ $\mu$ l)

Oxygen content	Compound	$n_{\mathrm{C,real}}/n_{\mathrm{H,real}}$	$n_{\rm C,calc.}/n_{\rm H,calc.}$	R.S.D. (%)	Δ (%)
ca. 1.5% (recommended)	C <sub>11</sub> H <sub>24</sub>	0.46	0.46	0.2	0
	$C_{12}H_{26}$	0.46	0.46	0.1	0
	$C_{13}H_{28}$	0.46	Reference	_	_
	$C_{14}H_{30}$	0.47	0.47	0.1	0
	$C_{15}H_{32}$	0.47	0.47	0.3	0
	$C_{16}H_{34}$	0.47	0.47	0.2	0

### Download English Version:

# https://daneshyari.com/en/article/9749177

Download Persian Version:

https://daneshyari.com/article/9749177

<u>Daneshyari.com</u>