

Short communication

High performance liquid chromatographic determination of azithromycin in serum using fluorescence detection and its application in human pharmacokinetic studies

Gholamreza Bahrami^{a,*}, Shahla Mirzaeei^b, Amir Kiani^a^a Medical Biology Research Center, Medical School, Kermanshah University of Medical Sciences, Kermanshah 20155, Iran^b School of Pharmacy, Kermanshah University of Medical Sciences, Kermanshah, Iran

Received 30 November 2004; accepted 5 March 2005

Available online 3 May 2005

Abstract

A fast and sensitive high-performance liquid chromatographic method for determination of azithromycin in human serum using fluorescence detection was developed. The drug and an internal standard (clarithromycin) were extracted from serum using *n*-hexan and subjected to pre-column derivatization with 9-fluorenylmethyl chloroformate as labeling agent. Analysis was performed on a phenyl packing material column with sodium phosphate buffer containing 2 ml/l triethylamine (pH 5.9) and methanol (29:71, v/v) as the mobile phase. The standard curve was linear over the range of 10–500 ng/ml of azithromycin in human serum. The means between-days precision were from 13.3% (for 10 ng/ml) to 2% (500 ng/ml) and the within-day precision from 11.9 to 1.7% determined on spiked samples. The accuracy of the method was 100.7–107.2% (between days) and 100.3–107.8% (within day). The limit of quantification was 10 ng/ml. This method was applied in a bioequivalence study of four different azithromycin preparations in 12 healthy volunteers.

© 2005 Elsevier B.V. All rights reserved.

Keywords: Reverse phase chromatography; HPLC; Azithromycin; Serum; Bioequivalence study; Macrolide antibiotics

1. Introduction

Azithromycin [9-de-oxo-9a-aza-9a-methyl-9a-homoerythromycin A dehydrate], is a macrolide antibibiotic. It is chemically related to erythromycin with enhanced spectrum and potency against some bacteria, longer elimination half life, superior stability in acid environment and higher tissue concentrations [1]. Low plasma concentration is achieved following administration of azithromycin, thus quantification methods of the drug in pharmacokinetic studies and monitoring of its efficacy needs to be sensitive and specific. The analysis of the drug is complicated, because azithromycin has only a weak UV absorbance in the wavelength range of less than 220 nm. High-performance liquid chromatography (HPLC) with UV detection has been used for analysis of the drug in in-vitro studies including bulk samples, raw materi-

als and acid degradation studies [2–4]. In biological samples several HPLC methods using electrochemical detection have been developed for determination of azithromycin [5–12]. A sensitive HPLC method using atmospheric pressure chemical ionization mass spectrometric detection has been reported for the analysis of azithromycin in plasma. This method requires only a sample volume of 50 µl [13]. Quantitative determination of azithromycin in human serum using 9-fluorenylmethyl chloroformate (FMOC-Cl) as a fluorescent labeling agent has been reported by Torano and Guchelaar [14]. In their method however, the limit of quantitation (LOQ) is 92 ng/ml which is not enough for human pharmacokinetic studies using low dosages. Also due to the simultaneous determination of all macrolides the retention time is 20 min, which is time consuming. Based on their method a new sensitive, rapid and simple assay for the determination of azithromycin in human serum has been developed. In the present method the sensitivity has been improved and the run time has been reduced. This method was applied for analysis of the drug in

* Corresponding author. Tel.: +98 831 8350197; fax: +98 831 8368410.
E-mail address: gbahrami@kums.ac.ir (G. Bahrami).

a bioequivalence study following oral administration of four different azithromycin preparations in 12 healthy volunteers.

2. Experimental

2.1. Chemicals

Azithromycin and clarithromycin hydrochloride (I.S.) were from Sigma (St. Louis, MO, USA) and kindly provided by Tehran Shimi pharmaceutical company (Tehran, Iran). Methanol (HPLC grade), hexan, boric acid, potassium chloride, potassium hydroxide, potassium dihydrogen phosphate, triethylamine, sodium carbonate, sodium bicarbonate, phosphoric acid and glycine were purchased from Merck (Darmstadt, Germany). FMOC-Cl was obtained from Sigma (St. Louis, MO, USA). All reagents used were of analytical grade except methanol which was HPLC grade. Water was glass-double distilled and further purified for HPLC with a Maxima purification system (USF ELGA, England).

2.2. Preparation of standard solutions

Stock solutions of azithromycin and the I.S. were prepared by dissolving the compounds in acetonitrile at concentrations of 1000 and 25 µg/ml, respectively. The azithromycin stock solution was further diluted with acetonitrile to working solutions ranging from 10 to 5000 ng/ml. A borate buffer (0.1 M) was prepared by dissolving 0.625 g of boric acid and 0.750 g of potassium chloride in 100 ml water and adjusting the pH to 7.4 with 0.2 M potassium hydroxide solution. A 500 µl/ml solution of FMOC-Cl was prepared in acetonitrile. Stock solution of glycine (4 mg/ml) was prepared in water. All solutions were stored at 4 °C and were stable for at least 4 weeks.

2.3. Chromatography

The HPLC system consisted of two pumps (LC-10AD), a column oven (CTO-10A), a spectrofluorometric detector (RF-551) operated at an excitation wavelength of 260 and emission wavelength of 315 nm, a degasser (DGU-3A) and a data processor (C-R4A) all from Shimadzu, Kyoto, Japan. The analysis was performed on a phenyl column (150 mm × 6 mm i.d.) which was packed with 5 µm particles (Shimpack-CLC-Phenyl) and operated at 62 °C. The mobile phase consisted of methanol–0.05 M sodium phosphate buffer (71/29, v/v) containing 2 ml/l triethylamine and adjusted to a pH of 5.9 with phosphoric acid. The eluent was filtered, degassed and pumped at a flow rate of 2.5 ml/min.

2.4. Sample preparation and derivatization

To 1 ml serum samples 100 µl of I.S., and 5 ml hexan were added. After mixing for 30 s on a vortex mixer and centrifugation (5 min at 6000 × g), the organic phase was removed and evaporated to dryness under a stream of nitrogen at 45 °C.

The residue was subjected to derivatization with FMOC-Cl as previously described [14] with some modifications. Briefly 160 µl FMOC-Cl (500 µl/ml in acetonitrile) and 40 µl of borate buffer (pH 7.4) were added and after vortex mixing for 10 s the samples were kept at 50 °C for 40 min. The reaction was stopped by adding 10 µl glycine (0.1 M) and, after 1 min, 20 µl of the reaction mixture was injected in to the chromatograph.

2.5. Calibration

An amount of 100 µl from each working solutions was evaporated in disposable glass tubes (16 mm × 100 mm) under a gentle stream of nitrogen at 50 °C. The residue was reconstituted in 1 ml drug-free human serum. The samples were then submitted to the procedures of extraction, derivatization and chromatographic analysis described above. Calibration curves (unweighted regression line) were obtained by linear least-squares regression analysis plotting of peak-area ratios (azithromycin/I.S.) versus the azithromycin concentrations.

2.6. Performance assessment

The recovery of azithromycin from serum was determined at concentration ranges of 10, 100 and 500 ng/ml by comparing peak areas obtained after derivatization of azithromycin extracted from serum with peak areas obtained after derivatization of the same amounts of unextracted azithromycin solutions in acetonitrile. The recovery of I.S. from serum was determined at a concentration of 25 ng/ml by the same method. Within-day variation was measured by assessing the different controls in replicates of six. Between-days variation was based on repeated analysis of the same concentration controls in ten analytical run performed on different days. The specificity of the method was investigated by the analysis of 12 human blank serum samples from different volunteers. These samples were pretreated according to the sample preparation procedure except from the addition of the I.S. The selectivity of the assay was evaluated by derivatization and analysis of a group of potentially co-administrated drugs with the azithromycin. The limit of detection was defined as the concentration of drug giving a signal to noise ratio of 4:1. The lower limit of quantification was defined as the lowest serum concentration of azithromycin quantified with a coefficient of variation of less than 20% (range recommended by the Conference Report on Bioanalytical Methods Validation [15]).

3. Results and discussion

3.1. Specificity and selectivity

Representative chromatograms of human blank serum and human blank serum spiked with azithromycin (10 ng/ml) and the I.S. are shown in Fig. 1A and B, respectively.

Download English Version:

<https://daneshyari.com/en/article/9753335>

Download Persian Version:

<https://daneshyari.com/article/9753335>

[Daneshyari.com](https://daneshyari.com)